

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091924\
 Data File : P0106692.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2024 15:05
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 05:17:35 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 09:57:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.452	3.714	410.0E6	137.9E6	54.656	51.616
2)	SA Decachlor...	10.198	8.738	246.7E6	106.7E6	49.904	53.350
Target Compounds							
26)	L6 AR-1254-1	6.470	5.605	146.1E6	71370269	560.589	506.684
27)	L6 AR-1254-2	6.688	5.751	203.7E6	62754386	522.663	492.547
28)	L6 AR-1254-3	7.053	6.157	200.3E6	96956699	513.844	481.585
29)	L6 AR-1254-4	7.337	6.384	144.8E6	59682474	498.346	479.603
30)	L6 AR-1254-5	7.757	6.802	150.1E6	87549838	493.066	488.803



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091924\
Data File : PO106692.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2024 15:05
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 05:17:35 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 09:57:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: -0.002 min
Response: 409993658
Conc: 54.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.714 min
Delta R.T.: -0.004 min
Response: 137878272
Conc: 51.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.198 min
Delta R.T.: 0.000 min
Response: 246666711
Conc: 49.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.738 min
Delta R.T.: -0.008 min
Response: 106712585
Conc: 53.35 ng/ml



#26 AR-1254-1


R.T.: 6.470 min
Delta R.T.: 0.001 min
Response: 146093951
Conc: 560.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.605 min
Delta R.T.: -0.003 min
Response: 71370269
Conc: 506.68 ng/ml




#27 AR-1254-2

R.T.: 6.688 min
Delta R.T.: 0.002 min
Response: 203654581
Conc: 522.66 ng/ml



#27 AR-1254-2


R.T.: 5.751 min
Delta R.T.: -0.004 min
Response: 62754386
Conc: 492.55 ng/ml



#28 AR-1254-3


R.T.: 7.053 min
Delta R.T.: 0.000 min
Response: 200347584
Conc: 513.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.157 min
Delta R.T.: -0.004 min
Response: 96956699
Conc: 481.58 ng/ml




#29 AR-1254-4

R.T.: 7.337 min
Delta R.T.: 0.000 min
Response: 144814070
Conc: 498.35 ng/ml



#29 AR-1254-4


R.T.: 6.384 min
Delta R.T.: -0.003 min
Response: 59682474
Conc: 479.60 ng/ml



#30 AR-1254-5

R.T.: 7.757 min
Delta R.T.: 0.000 min
Response: 150060073
Conc: 493.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.802 min
Delta R.T.: -0.004 min
Response: 87549838
Conc: 488.80 ng/ml