

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092023\
 Data File : P0098194.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2023 01:59
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/21/2023
 Supervised By :Ankita Jodhani 09/21/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 02:13:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 07:12:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.494	3.636	127.8E6	61513354	42.723	55.407 #
2)	SA Decachlor...	10.336	8.653	76447601	49817032	43.790m	45.220
Target Compounds							
3)	L1 AR-1016-1	5.674	4.723	40190559	21825278	437.033	486.951
4)	L1 AR-1016-2	5.697	4.742	57396476	28029640	432.434	471.609m
5)	L1 AR-1016-3	5.759	4.918	36278967	16262800	428.370	503.094
6)	L1 AR-1016-4	5.858	4.960	28839878	12463779	433.711	473.890
7)	L1 AR-1016-5	6.156	5.174	28077519	16676280	450.794	483.577m
31)	L7 AR-1260-1	7.291	6.210	51528838	31895147	458.626	460.359
32)	L7 AR-1260-2	7.548	6.399	57164188	36722804	436.993	448.502
33)	L7 AR-1260-3	7.910	6.552	40226277	34559872	417.109	446.157
34)	L7 AR-1260-4	8.138	7.025	44707012	27927039	407.521	449.165
35)	L7 AR-1260-5	8.465	7.267	86779712	64447641	422.724	454.343

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092023\
Data File : PO098194.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2023 01:59
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/21/2023
Supervised By :Ankita Jodhani 09/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 02:13:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 07:12:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

