

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092024\
 Data File : P0106762.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2024 05:47
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/23/2024
 Supervised By :Ankita Jodhani 09/23/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 07:11:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 09:57:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.451	3.712	359.4E6	129.0E6	47.914	48.306
2)	SA Decachlor...	10.194	8.736	217.4E6	92200500	43.973	46.095
Target Compounds							
3)	L1 AR-1016-1	5.615	4.799	122.5E6	43771423	479.720	459.529
4)	L1 AR-1016-2	5.638	4.819	171.9E6	59134219	482.833	462.668
5)	L1 AR-1016-3	5.700	4.996	102.4E6	31232339	469.633	452.757
6)	L1 AR-1016-4	5.799	5.036	84644944	22383204	477.245	421.470
7)	L1 AR-1016-5	6.092	5.250	80955650	31499011	469.063	440.942m
31)	L7 AR-1260-1	7.217	6.285	137.3E6	60016815	486.605	464.079
32)	L7 AR-1260-2	7.472	6.472	146.9E6	75816544	440.246	477.708
33)	L7 AR-1260-3	7.831	6.627	86722651	70585028	384.218	457.615
34)	L7 AR-1260-4	8.056	7.098	130.7E6	51183650	472.546	481.427
35)	L7 AR-1260-5	8.375	7.338	237.0E6	128.0E6	480.125	497.752

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092024\
Data File : PO106762.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2024 05:47
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/23/2024
Supervised By :Ankita Jodhani 09/23/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 07:11:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 09:57:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

