

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092120\
 Data File : P0071512.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Sep 2020 17:45
 Operator : DD\AJ
 Sample : PB131750BL
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB131750BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 05:08:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 09 16:04:54 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.341	3.526	1520249	807717	21.452	20.689
2) SA Decachlor...	9.952	8.702	1844516	1022693	19.779	18.566
Target Compounds						
24) L5 AR-1248-4	6.547f	0.000	65250	0	17.956	N.D. #
26) L6 AR-1254-1	6.547	0.000	65250	0	18.598	N.D. #
28) L6 AR-1254-3	7.168f	0.000	214597	0	36.084	N.D. #
29) L6 AR-1254-4	7.466f	0.000	177138	0	30.493	N.D. #
39) L8 AR-1262-4	8.807	0.000	171003	0	26.894	N.D. #
42) L9 AR-1268-2	8.807	0.000	171003	0	12.232	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092120\
Data File : P0071512.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Sep 2020 17:45
Operator : DD\AJ
Sample : PB131750BL
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB131750BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 05:08:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090920.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 09 16:04:54 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.341 min
Delta R.T.: -0.003 min
Response: 1520249
Conc: 21.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB131750BL




#1 Tetrachloro-m-xylene

R.T.: 3.526 min
Delta R.T.: -0.002 min
Response: 807717
Conc: 20.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.952 min
Delta R.T.: -0.007 min
Response: 1844516
Conc: 19.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.702 min
Delta R.T.: -0.006 min
Response: 1022693
Conc: 18.57 ng/ml



#24 AR-1248-4


R.T.: 6.547 min
Delta R.T.: -0.025 min
Response: 65250
Conc: 17.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB131750BL




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.233 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.547 min
Delta R.T.: 0.010 min
Response: 65250
Conc: 18.60 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.607 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.168 min
Delta R.T.: 0.026 min
Response: 214597
Conc: 36.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB131750BL




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.177 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 7.466 min
Delta R.T.: 0.029 min
Response: 177138
Conc: 30.49 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 6.418 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.807 min
Delta R.T.: -0.010 min
Response: 171003
Conc: 26.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB131750BL




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.732 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 8.807 min
Delta R.T.: -0.012 min
Response: 171003
Conc: 12.23 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 7.735 min
Response: 0
Conc: N.D.