

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092123\
 Data File : P0098219.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2023 14:55
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/22/2023
 Supervised By :Ankita Jodhani 09/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 15:41:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 07:12:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.493	3.638	129.8E6	61632588	43.413m	55.514m#
2) SA Decachlor...	10.334	8.654	77655779	51122996	44.482	46.406
Target Compounds						
3) L1 AR-1016-1	5.672	4.724	40075703	21819816	435.784	486.829
4) L1 AR-1016-2	5.695	4.743	57908916	28273310	436.295	475.709m
5) L1 AR-1016-3	5.758	4.919	35546330	15763056	419.719	487.634m
6) L1 AR-1016-4	5.857	4.961	28810187	9647134	433.265	366.797m
7) L1 AR-1016-5	6.154	5.175	28484504	16627675	457.329	482.168m
31) L7 AR-1260-1	7.288	6.210	50113277	31944807	446.027m	461.076m
32) L7 AR-1260-2	7.546	6.399	57538145	38452025	439.852m	469.621
33) L7 AR-1260-3	7.908	6.553	43845399	35636303	454.636	460.054
34) L7 AR-1260-4	8.136	7.026	47525994	28365519	433.217	456.217
35) L7 AR-1260-5	8.463	7.267	89056566	66191102	433.815	466.634

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092123\
Data File : PO098219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2023 14:55
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/22/2023
Supervised By :Ankita Jodhani 09/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 15:41:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 07:12:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

