

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092225\
 Data File : P0113905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 12:06
 Operator : YP/AJ
 Sample : Q3147-10 10X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S-10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 12:22:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 16 02:49:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.660	3.650	12540921	8082662	1.422	1.513
2)	SA Decachlor...	8.669	8.612	9860551	3608728	1.474	1.729
Target Compounds							
31)	L7 AR-1260-1	6.227	6.193	560.4E6	260.1E6	1143.299	1035.044
32)	L7 AR-1260-2	6.417	6.382	811.6E6	315.7E6	1059.731	955.014
33)	L7 AR-1260-3	6.781	6.532	676.6E6	259.0E6	1005.595	923.968
34)	L7 AR-1260-4	7.040	7.001	479.9E6	171.1E6	1108.963	895.024
35)	L7 AR-1260-5	7.284	7.244	1351.5E6	398.6E6	1193.034	936.444



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092225\
Data File : PO113905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 12:06
Operator : YP/AJ
Sample : Q3147-10 10X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S-10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 12:22:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 16 02:49:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.660 min
Delta R.T.: 0.000 min
Response: 12540921
Conc: 1.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-10




#1 Tetrachloro-m-xylene

R.T.: 3.650 min
Delta R.T.: -0.001 min
Response: 8082662
Conc: 1.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.669 min
Delta R.T.: -0.001 min
Response: 9860551
Conc: 1.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.612 min
Delta R.T.: -0.002 min
Response: 3608728
Conc: 1.73 ng/ml



#31 AR-1260-1


R.T.: 6.227 min
Delta R.T.: -0.002 min
Response: 560446903
Conc: 1143.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-10




#31 AR-1260-1

R.T.: 6.193 min
Delta R.T.: -0.003 min
Response: 260095569
Conc: 1035.04 ng/ml




#32 AR-1260-2

R.T.: 6.417 min
Delta R.T.: -0.001 min
Response: 811584162
Conc: 1059.73 ng/ml



#32 AR-1260-2


R.T.: 6.382 min
Delta R.T.: -0.002 min
Response: 315735320
Conc: 955.01 ng/ml



#33 AR-1260-3


R.T.: 6.781 min
Delta R.T.: -0.001 min
Response: 676576446
Conc: 1005.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-10




#33 AR-1260-3

R.T.: 6.532 min
Delta R.T.: -0.003 min
Response: 259041180
Conc: 923.97 ng/ml




#34 AR-1260-4

R.T.: 7.040 min
Delta R.T.: 0.000 min
Response: 479887996
Conc: 1108.96 ng/ml



#34 AR-1260-4


R.T.: 7.001 min
Delta R.T.: -0.002 min
Response: 171138667
Conc: 895.02 ng/ml



#35 AR-1260-5

R.T.: 7.284 min
Delta R.T.: -0.001 min
Response: 1351477817
Conc: 1193.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-10



#35 AR-1260-5

R.T.: 7.244 min
Delta R.T.: -0.002 min
Response: 398624497
Conc: 936.44 ng/ml