

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092320\
 Data File : P0071590.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Sep 2020 19:01
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 19:18:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 23 08:07:44 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.342	3.525	4742317	2319854	55.263	56.327
2)	SA Decachlor...	9.954	8.702	6029741	3038033	52.484	49.980
Target Compounds							
3)	L1 AR-1016-1	5.644	4.748	1990243	954477	546.487	540.971
4)	L1 AR-1016-2	5.667	4.766	2877420	1361777	544.982	548.668
5)	L1 AR-1016-3	5.729	4.950	1701024	735746	539.075	546.253
6)	L1 AR-1016-4	5.836	5.007	1446001	638408	539.823	546.330
7)	L1 AR-1016-5	6.145	5.227	1420727	763719	549.039	540.198
31)	L7 AR-1260-1	7.301	6.307	3089525	1443175	543.707	521.484
32)	L7 AR-1260-2	7.568	6.508	4408674	1848658	507.829	523.020
33)	L7 AR-1260-3	7.924	6.655	3786053	1687579	525.809	524.955
34)	L7 AR-1260-4	8.152	7.130	3572782	1513116	539.263	524.445
35)	L7 AR-1260-5	8.467	7.382	7892828	3897177	523.428	526.877



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092320\
Data File : PO071590.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Sep 2020 19:01
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 19:18:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092320.M
Quant Title : GC EXTRCTABLES
QLast Update : Wed Sep 23 08:07:44 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.342 min
Delta R.T.: 0.000 min
Response: 4742317
Conc: 55.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.525 min
Delta R.T.: 0.000 min
Response: 2319854
Conc: 56.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.954 min
Delta R.T.: -0.003 min
Response: 6029741
Conc: 52.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.702 min
Delta R.T.: -0.002 min
Response: 3038033
Conc: 49.98 ng/ml



#3 AR-1016-1


R.T.: 5.644 min
Delta R.T.: -0.001 min
Response: 1990243
Conc: 546.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.748 min
Delta R.T.: 0.000 min
Response: 954477
Conc: 540.97 ng/ml




#4 AR-1016-2

R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 2877420
Conc: 544.98 ng/ml



#4 AR-1016-2


R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 1361777
Conc: 548.67 ng/ml



#5 AR-1016-3


R.T.: 5.729 min
Delta R.T.: 0.000 min
Response: 1701024
Conc: 539.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.950 min
Delta R.T.: -0.001 min
Response: 735746
Conc: 546.25 ng/ml




#6 AR-1016-4

R.T.: 5.836 min
Delta R.T.: 0.000 min
Response: 1446001
Conc: 539.82 ng/ml



#6 AR-1016-4


R.T.: 5.007 min
Delta R.T.: -0.001 min
Response: 638408
Conc: 546.33 ng/ml



#7 AR-1016-5


R.T.: 6.145 min
Delta R.T.: 0.000 min
Response: 1420727
Conc: 549.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.227 min
Delta R.T.: -0.002 min
Response: 763719
Conc: 540.20 ng/ml




#31 AR-1260-1

R.T.: 7.301 min
Delta R.T.: -0.001 min
Response: 3089525
Conc: 543.71 ng/ml



#31 AR-1260-1


R.T.: 6.307 min
Delta R.T.: -0.002 min
Response: 1443175
Conc: 521.48 ng/ml



#32 AR-1260-2


R.T.: 7.568 min
Delta R.T.: -0.002 min
Response: 4408674
Conc: 507.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.508 min
Delta R.T.: -0.002 min
Response: 1848658
Conc: 523.02 ng/ml




#33 AR-1260-3

R.T.: 7.924 min
Delta R.T.: -0.003 min
Response: 3786053
Conc: 525.81 ng/ml



#33 AR-1260-3


R.T.: 6.655 min
Delta R.T.: -0.002 min
Response: 1687579
Conc: 524.95 ng/ml



#34 AR-1260-4


R.T.: 8.152 min
Delta R.T.: -0.002 min
Response: 3572782
Conc: 539.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.130 min
Delta R.T.: -0.002 min
Response: 1513116
Conc: 524.45 ng/ml



#35 AR-1260-5

R.T.: 8.467 min
Delta R.T.: -0.001 min
Response: 7892828
Conc: 523.43 ng/ml



#35 AR-1260-5

R.T.: 7.382 min
Delta R.T.: -0.003 min
Response: 3897177
Conc: 526.88 ng/ml