

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092321\
 Data File : P0081451.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Sep 2021 16:25
 Operator : AJ\MA
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 17:00:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 23 16:59:08 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.869	3.987	2376986	736213	24.700	25.816
2)	SA Decachlor...	10.882	9.335	2044088	668896	25.850	26.652
Target Compounds							
21)	L5 AR-1248-1	6.196	5.256	519182	155532	257.831	257.107
22)	L5 AR-1248-2	6.492	5.522	685294	239504	260.440	268.266
23)	L5 AR-1248-3	6.710	5.566	813302	237606	261.446	267.558
24)	L5 AR-1248-4	7.141	5.753	886384	270699	261.036	266.705
25)	L5 AR-1248-5	7.180	6.179	875829	256242	262.623	266.511



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092321\
Data File : PO081451.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Sep 2021 16:25
Operator : AJ\MA
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:00:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092321.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 23 16:59:08 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.869 min
Delta R.T.: -0.001 min
Response: 2376986
Conc: 24.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.987 min
Delta R.T.: 0.000 min
Response: 736213
Conc: 25.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.882 min
Delta R.T.: 0.000 min
Response: 2044088
Conc: 25.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.335 min
Delta R.T.: 0.000 min
Response: 668896
Conc: 26.65 ng/ml



#21 AR-1248-1


R.T.: 6.196 min
Delta R.T.: 0.000 min
Response: 519182
Conc: 257.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 155532
Conc: 257.11 ng/ml




#22 AR-1248-2

R.T.: 6.492 min
Delta R.T.: 0.000 min
Response: 685294
Conc: 260.44 ng/ml



#22 AR-1248-2


R.T.: 5.522 min
Delta R.T.: 0.000 min
Response: 239504
Conc: 268.27 ng/ml



#23 AR-1248-3


R.T.: 6.710 min
Delta R.T.: 0.000 min
Response: 813302
Conc: 261.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 237606
Conc: 267.56 ng/ml




#24 AR-1248-4

R.T.: 7.141 min
Delta R.T.: 0.000 min
Response: 886384
Conc: 261.04 ng/ml



#24 AR-1248-4


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 270699
Conc: 266.71 ng/ml



#25 AR-1248-5

R.T.: 7.180 min
Delta R.T.: 0.000 min
Response: 875829
Conc: 262.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 6.179 min
Delta R.T.: 0.000 min
Response: 256242
Conc: 266.51 ng/ml