

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092324\
 Data File : P0106778.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2024 11:20
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:43:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 09:57:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.449	3.709	370.4E6	151.9E6	49.379	56.877
2)	SA Decachlor...	10.193	8.733	243.1E6	111.9E6	49.183	55.967
Target Compounds							
26)	L6 AR-1254-1	6.468	5.601	129.3E6	69746570	495.993	495.157
27)	L6 AR-1254-2	6.686	5.748	185.0E6	62058880	474.813	487.088
28)	L6 AR-1254-3	7.052	6.154	182.1E6	98316321	466.951	488.338
29)	L6 AR-1254-4	7.336	6.381	143.2E6	60136613	492.906	483.252
30)	L6 AR-1254-5	7.754	6.799	147.7E6	88571830	485.339	494.509



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092324\
Data File : PO106778.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 11:20
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 09:57:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.005 min
Response: 370412672
Conc: 49.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.709 min
Delta R.T.: -0.009 min
Response: 151931470
Conc: 56.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.193 min
Delta R.T.: -0.004 min
Response: 243102639
Conc: 49.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.733 min
Delta R.T.: -0.014 min
Response: 111947394
Conc: 55.97 ng/ml



#26 AR-1254-1


R.T.: 6.468 min
Delta R.T.: -0.001 min
Response: 129259682
Conc: 495.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.601 min
Delta R.T.: -0.007 min
Response: 69746570
Conc: 495.16 ng/ml




#27 AR-1254-2

R.T.: 6.686 min
Delta R.T.: 0.000 min
Response: 185010145
Conc: 474.81 ng/ml



#27 AR-1254-2


R.T.: 5.748 min
Delta R.T.: -0.007 min
Response: 62058880
Conc: 487.09 ng/ml



#28 AR-1254-3


R.T.: 7.052 min
Delta R.T.: 0.000 min
Response: 182064190
Conc: 466.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.154 min
Delta R.T.: -0.007 min
Response: 98316321
Conc: 488.34 ng/ml




#29 AR-1254-4

R.T.: 7.336 min
Delta R.T.: 0.000 min
Response: 143233375
Conc: 492.91 ng/ml



#29 AR-1254-4


R.T.: 6.381 min
Delta R.T.: -0.006 min
Response: 60136613
Conc: 483.25 ng/ml



#30 AR-1254-5

R.T.: 7.754 min
Delta R.T.: -0.002 min
Response: 147708423
Conc: 485.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.799 min
Delta R.T.: -0.008 min
Response: 88571830
Conc: 494.51 ng/ml