

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092419\
 Data File : P0061313.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Sep 2019 11:48
 Operator : SM/AJ
 Sample : K4997-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 WC-08-091919

Manual Integrations
 APPROVED

Ankita
 9/25/2019 8:45:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 02:35:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 14 00:06:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.369	3.529	928617	704637	25.240	25.236
2) SA Decachlor...	10.036	8.403	1103004	791185	24.006	22.572m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092419\
Data File : P0061313.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Sep 2019 11:48
Operator : SM/AJ
Sample : K4997-01
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
WC-08-091919


Manual Integrations
APPROVED

Ankita
9/25/2019 8:45:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 02:35:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.369 min
Delta R.T.: 0.000 min
Response: 928617
Conc: 25.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
WC-08-091919


Manual Integrations
APPROVED

Ankita
9/25/2019 8:45:59 AM




#1 Tetrachloro-m-xylene

R.T.: 3.529 min
Delta R.T.: -0.003 min
Response: 704637
Conc: 25.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.036 min
Delta R.T.: 0.008 min
Response: 1103004
Conc: 24.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.403 min
Delta R.T.: -0.010 min
Response: 791185
Conc: 22.57 ng/ml m