

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092421\
 Data File : P0081474.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Sep 2021 11:07
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 01:12:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 02:21:54 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.869	3.987	5079538	1460103	56.961	54.805
2) SA Decachlor...	10.879	9.333	4286153	1315141	57.542	56.632
Target Compounds						
26) L6 AR-1254-1	7.108	6.134	2036451	802010	567.973	523.760
27) L6 AR-1254-2	7.338	6.295	2880310	733222	522.372	538.920
28) L6 AR-1254-3	7.721	6.719	3073856	1056259	544.766	531.005
29) L6 AR-1254-4	8.018	6.960	2246624	631323	546.348	505.610
30) L6 AR-1254-5	8.447	7.392	2460599	988230	533.824	528.629



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092421\
Data File : PO081474.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Sep 2021 11:07
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 01:12:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 24 02:21:54 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.869 min
Delta R.T.: -0.002 min
Response: 5079538
Conc: 56.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.987 min
Delta R.T.: -0.002 min
Response: 1460103
Conc: 54.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.879 min
Delta R.T.: -0.007 min
Response: 4286153
Conc: 57.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.333 min
Delta R.T.: -0.004 min
Response: 1315141
Conc: 56.63 ng/ml



#26 AR-1254-1


R.T.: 7.108 min
Delta R.T.: -0.003 min
Response: 2036451
Conc: 567.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.134 min
Delta R.T.: -0.002 min
Response: 802010
Conc: 523.76 ng/ml




#27 AR-1254-2

R.T.: 7.338 min
Delta R.T.: -0.003 min
Response: 2880310
Conc: 522.37 ng/ml



#27 AR-1254-2


R.T.: 6.295 min
Delta R.T.: -0.002 min
Response: 733222
Conc: 538.92 ng/ml



#28 AR-1254-3


R.T.: 7.721 min
Delta R.T.: -0.003 min
Response: 3073856
Conc: 544.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.719 min
Delta R.T.: -0.002 min
Response: 1056259
Conc: 531.00 ng/ml




#29 AR-1254-4

R.T.: 8.018 min
Delta R.T.: -0.003 min
Response: 2246624
Conc: 546.35 ng/ml



#29 AR-1254-4


R.T.: 6.960 min
Delta R.T.: -0.002 min
Response: 631323
Conc: 505.61 ng/ml



#30 AR-1254-5

R.T.: 8.447 min
Delta R.T.: -0.003 min
Response: 2460599
Conc: 533.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.392 min
Delta R.T.: -0.002 min
Response: 988230
Conc: 528.63 ng/ml