

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092425\
 Data File : P0114003.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2025 19:16
 Operator : YP/AJ
 Sample : Q3169-05DL 50X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CT#1DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/25/2025
 Supervised By :mohammad ahmed 09/26/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 08:05:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 16 02:49:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds								
26)	L6	AR-1254-1	5.544	5.516	319.3E6	168.1E6	585.673	555.882
27)	L6	AR-1254-2	5.693	5.664	300.7E6	149.8E6	591.363	567.168
28)	L6	AR-1254-3	6.095	6.062	389.3E6	197.5E6	497.791	509.856
29)	L6	AR-1254-4	6.324	6.290	272.6E6	124.2E6	498.406	523.038
30)	L6	AR-1254-5	6.742	6.705	343.2E6	147.2E6	461.057	479.672

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092425\
Data File : PO114003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2025 19:16
Operator : YP/AJ
Sample : Q3169-05DL 50X
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CT#1DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/25/2025
Supervised By :mohammad ahmed 09/26/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 08:05:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 16 02:49:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.660 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
CT#1DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/25/2025
Supervised By :mohammad ahmed 09/26/2025




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.651 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.670 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.614 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 5.544 min
Delta R.T.: -0.002 min
Response: 319320100
Conc: 585.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
CT#1DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/25/2025
Supervised By :mohammad ahmed 09/26/2025




#26 AR-1254-1

R.T.: 5.516 min
Delta R.T.: -0.003 min
Response: 168097634
Conc: 555.88 ng/ml




#27 AR-1254-2

R.T.: 5.693 min
Delta R.T.: -0.002 min
Response: 300671865
Conc: 591.36 ng/ml



#27 AR-1254-2

R.T.: 5.664 min
Delta R.T.: -0.004 min
Response: 149777134
Conc: 567.17 ng/ml




#28 AR-1254-3

R.T.: 6.095 min
Delta R.T.: -0.003 min
Response: 389324763
Conc: 497.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
CT#1DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/25/2025
Supervised By :mohammad ahmed 09/26/2025




#28 AR-1254-3

R.T.: 6.062 min
Delta R.T.: -0.004 min
Response: 197476130
Conc: 509.86 ng/ml




#29 AR-1254-4

R.T.: 6.324 min
Delta R.T.: -0.004 min
Response: 272555129
Conc: 498.41 ng/ml



#29 AR-1254-4

R.T.: 6.290 min
Delta R.T.: -0.004 min
Response: 124165091
Conc: 523.04 ng/ml




#30 AR-1254-5

R.T.: 6.742 min
Delta R.T.: -0.002 min
Response: 343181700
Conc: 461.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
CT#1DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/25/2025
Supervised By :mohammad ahmed 09/26/2025



#30 AR-1254-5

R.T.: 6.705 min
Delta R.T.: -0.003 min
Response: 147246021
Conc: 479.67 ng/ml