

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092519\
 Data File : P0061343.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Sep 2019 10:23
 Operator : SM/AJ
 Sample : K5008-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-1

Manual Integrations
 APPROVED

Ankita
 9/26/2019 5:02:33 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 01:18:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 14 00:06:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.367	3.528	1001392	626721	27.218m	22.445m
2) SA Decachlor...	10.030	8.400	896215	603070	19.506	17.205m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092519\
Data File : P0061343.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Sep 2019 10:23
Operator : SM/AJ
Sample : K5008-01
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SB-1


Manual Integrations
APPROVED

Ankita
9/26/2019 5:02:33 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 01:18:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.367 min
Delta R.T.: -0.003 min
Response: 1001392
Conc: 27.22 ng/ml m

Instrument :
ECD_O
ClientSampleId :
SB-1


Manual Integrations
APPROVED

Ankita
9/26/2019 5:02:33 PM




#1 Tetrachloro-m-xylene

R.T.: 3.528 min
Delta R.T.: -0.003 min
Response: 626721
Conc: 22.45 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.030 min
Delta R.T.: 0.003 min
Response: 896215
Conc: 19.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.400 min
Delta R.T.: -0.013 min
Response: 603070
Conc: 17.21 ng/ml m