

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092523\
 Data File : P0098232.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2023 18:11
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 22:41:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Sep 25 22:38:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.483	3.642	184.5E6	74037904	94.100	94.306
2) SA Decachlor...	10.299	8.663	87136703	39292191	96.213	94.591
Target Compounds						
16) L4 AR-1242-1	5.658	4.728	37567612	16930603	902.182	910.944
17) L4 AR-1242-2	5.680	4.747	59535628	24551081	938.381	929.258
18) L4 AR-1242-3	5.743	4.923	36525569	13148150	923.927	925.877
19) L4 AR-1242-4	5.842	5.007	27728794	13885292	932.764	908.211
20) L4 AR-1242-5	6.582	5.532	27397499	14372429	978.917	921.133



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092523\
Data File : PO098232.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2023 18:11
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 22:41:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092523.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Sep 25 22:38:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.483 min
Delta R.T.: 0.000 min
Response: 184531066
Conc: 94.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.642 min
Delta R.T.: 0.000 min
Response: 74037904
Conc: 94.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.299 min
Delta R.T.: 0.000 min
Response: 87136703
Conc: 96.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.663 min
Delta R.T.: 0.000 min
Response: 39292191
Conc: 94.59 ng/ml



#16 AR-1242-1


R.T.: 5.658 min
Delta R.T.: 0.000 min
Response: 37567612
Conc: 902.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#16 AR-1242-1

R.T.: 4.728 min
Delta R.T.: 0.000 min
Response: 16930603
Conc: 910.94 ng/ml




#17 AR-1242-2

R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 59535628
Conc: 938.38 ng/ml



#17 AR-1242-2


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 24551081
Conc: 929.26 ng/ml



#18 AR-1242-3


R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 36525569
Conc: 923.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#18 AR-1242-3

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 13148150
Conc: 925.88 ng/ml




#19 AR-1242-4

R.T.: 5.842 min
Delta R.T.: 0.000 min
Response: 27728794
Conc: 932.76 ng/ml



#19 AR-1242-4


R.T.: 5.007 min
Delta R.T.: 0.000 min
Response: 13885292
Conc: 908.21 ng/ml



#20 AR-1242-5

R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 27397499
Conc: 978.92 ng/ml

Instrument :
ECD_O
ClientSampleId :



#20 AR-1242-5

R.T.: 5.532 min
Delta R.T.: 0.000 min
Response: 14372429
Conc: 921.13 ng/ml