

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092525\
 Data File : P0114056.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2025 01:55
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 05:13:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 16 02:49:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.658	3.648	474.5E6	270.0E6	53.805	50.561
2)	SA Decachlor...	8.667	8.612	363.9E6	106.1E6	54.393	50.806
Target Compounds							
26)	L6 AR-1254-1	5.544	5.515	310.2E6	155.1E6	568.935	512.898
27)	L6 AR-1254-2	5.693	5.662	285.0E6	134.6E6	560.628	509.619
28)	L6 AR-1254-3	6.095	6.062	416.6E6	196.3E6	532.691	506.748
29)	L6 AR-1254-4	6.325	6.290	298.6E6	122.2E6	546.059	514.818
30)	L6 AR-1254-5	6.743	6.705	385.6E6	149.3E6	518.084	486.330



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092525\
Data File : PO114056.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2025 01:55
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 05:13:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 16 02:49:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.658 min
Delta R.T.: -0.002 min
Response: 474540831
Conc: 53.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: -0.003 min
Response: 270048183
Conc: 50.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.667 min
Delta R.T.: -0.002 min
Response: 363906170
Conc: 54.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.612 min
Delta R.T.: -0.002 min
Response: 106069546
Conc: 50.81 ng/ml



#26 AR-1254-1


R.T.: 5.544 min
Delta R.T.: -0.002 min
Response: 310194030
Conc: 568.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.515 min
Delta R.T.: -0.004 min
Response: 155099347
Conc: 512.90 ng/ml




#27 AR-1254-2

R.T.: 5.693 min
Delta R.T.: -0.002 min
Response: 285044852
Conc: 560.63 ng/ml



#27 AR-1254-2


R.T.: 5.662 min
Delta R.T.: -0.005 min
Response: 134579708
Conc: 509.62 ng/ml



#28 AR-1254-3


R.T.: 6.095 min
Delta R.T.: -0.003 min
Response: 416620400
Conc: 532.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.062 min
Delta R.T.: -0.004 min
Response: 196272598
Conc: 506.75 ng/ml




#29 AR-1254-4

R.T.: 6.325 min
Delta R.T.: -0.002 min
Response: 298614247
Conc: 546.06 ng/ml



#29 AR-1254-4


R.T.: 6.290 min
Delta R.T.: -0.004 min
Response: 122213707
Conc: 514.82 ng/ml



#30 AR-1254-5

R.T.: 6.743 min
Delta R.T.: -0.002 min
Response: 385628892
Conc: 518.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.705 min
Delta R.T.: -0.003 min
Response: 149289721
Conc: 486.33 ng/ml