

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092623\
 Data File : P0098268.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2023 14:42
 Operator : YP/AJ
 Sample : 04599-03 10X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 KMA809L-2-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 15:30:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 26 09:03:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.483	3.640	3763212	1601552	1.971	2.110
2) SA Decachlor...	10.288	8.657	7115268	864355	8.852	2.245 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092623\
Data File : PO098268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2023 14:42
Operator : YP/AJ
Sample : 04599-03 10X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
KMA809L-2-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 15:30:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092523.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 26 09:03:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.483 min
Delta R.T.: 0.000 min
Response: 3763212
Conc: 1.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
KMA809L-2-1




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: -0.002 min
Response: 1601552
Conc: 2.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.288 min
Delta R.T.: -0.014 min
Response: 7115268
Conc: 8.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.657 min
Delta R.T.: -0.007 min
Response: 864355
Conc: 2.24 ng/ml