

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092721\
 Data File : P0081593.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Sep 2021 19:51
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 03:51:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 02:21:54 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.871	3.987	5099978	1426084	57.190	53.528
2)	SA Decachlor...	10.887	9.334	3746543	1184488	50.298	51.005
Target Compounds							
26)	L6 AR-1254-1	7.112	6.135	1882605	751266	525.065	490.621
27)	L6 AR-1254-2	7.342	6.295	2809551	665112	509.539	488.858
28)	L6 AR-1254-3	7.725	6.720	2849071	956562	504.928	480.885
29)	L6 AR-1254-4	8.021	6.961	2084274	604457	506.867	484.095
30)	L6 AR-1254-5	8.451	7.393	2275426	909124	493.651	486.313



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092721\
Data File : PO081593.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Sep 2021 19:51
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 03:51:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 24 02:21:54 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.871 min
Delta R.T.: 0.000 min
Response: 5099978
Conc: 57.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.987 min
Delta R.T.: -0.002 min
Response: 1426084
Conc: 53.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.887 min
Delta R.T.: 0.000 min
Response: 3746543
Conc: 50.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.334 min
Delta R.T.: -0.003 min
Response: 1184488
Conc: 51.01 ng/ml



#26 AR-1254-1


R.T.: 7.112 min
Delta R.T.: 0.000 min
Response: 1882605
Conc: 525.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.135 min
Delta R.T.: 0.000 min
Response: 751266
Conc: 490.62 ng/ml




#27 AR-1254-2

R.T.: 7.342 min
Delta R.T.: 0.001 min
Response: 2809551
Conc: 509.54 ng/ml



#27 AR-1254-2


R.T.: 6.295 min
Delta R.T.: -0.001 min
Response: 665112
Conc: 488.86 ng/ml



#28 AR-1254-3


R.T.: 7.725 min
Delta R.T.: 0.000 min
Response: 2849071
Conc: 504.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.720 min
Delta R.T.: -0.001 min
Response: 956562
Conc: 480.88 ng/ml




#29 AR-1254-4

R.T.: 8.021 min
Delta R.T.: 0.000 min
Response: 2084274
Conc: 506.87 ng/ml



#29 AR-1254-4


R.T.: 6.961 min
Delta R.T.: 0.000 min
Response: 604457
Conc: 484.09 ng/ml



#30 AR-1254-5

R.T.: 8.451 min
Delta R.T.: 0.001 min
Response: 2275426
Conc: 493.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.393 min
Delta R.T.: -0.001 min
Response: 909124
Conc: 486.31 ng/ml