

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092821\
 Data File : P0081622.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Sep 2021 15:15
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 15:46:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 02:21:54 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.869	3.985	2493832	693526	27.965	26.031
2)	SA Decachlor...	10.880	9.331	2079159	655953	27.913	28.246
Target Compounds							
8)	L2 AR-1221-1	5.122	0.000	57232	0	52.301	N.D. #
9)	L2 AR-1221-2	5.222	0.000	36411	0	46.180	N.D. #
24)	L5 AR-1248-4	7.116f	5.774f	102491	-3500	29.758	N.D. #
26)	L6 AR-1254-1	7.116	0.000	102491	0	28.585	N.D. #
28)	L6 AR-1254-3	7.738	0.000	49044	0	8.692	N.D. #
32)	L7 AR-1260-2	8.131f	0.000	53011	0	10.213	N.D. #
33)	L7 AR-1260-3	0.000	7.205	0	1137	N.D.	0.684 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092821\
Data File : P0081622.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Sep 2021 15:15
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 15:46:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 24 02:21:54 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.869 min
Delta R.T.: -0.002 min
Response: 2493832
Conc: 27.97 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.985 min
Delta R.T.: -0.003 min
Response: 693526
Conc: 26.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.880 min
Delta R.T.: -0.006 min
Response: 2079159
Conc: 27.91 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.331 min
Delta R.T.: -0.006 min
Response: 655953
Conc: 28.25 ng/ml



#8 AR-1221-1


R.T.: 5.122 min
Delta R.T.: 0.000 min
Response: 57232
Conc: 52.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.248 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 36411
Conc: 46.18 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.348 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 7.116 min
Delta R.T.: -0.024 min
Response: 102491
Conc: 29.76 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 5.774 min
Delta R.T.: 0.022 min
Response: -3500
Conc: N.D.




#26 AR-1254-1

R.T.: 7.116 min
Delta R.T.: 0.005 min
Response: 102491
Conc: 28.59 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T.: 6.136 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.738 min
Delta R.T.: 0.014 min
Response: 49044
Conc: 8.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.721 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 8.131 min
Delta R.T.: -0.027 min
Response: 53011
Conc: 10.21 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 7.057 min
Response: 0
Conc: N.D.



#33 AR-1260-3

R.T.: 0.000 min
Exp R.T.: 8.523 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#33 AR-1260-3

R.T.: 7.205 min
Delta R.T.: -0.008 min
Response: 1137
Conc: 0.68 ng/ml