

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092923\
 Data File : P0098416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2023 20:12
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 00:46:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.484	3.637	88724586	36919333	46.206	47.402
2) SA Decachlor...	10.299	8.648	58305044	26530457	48.033	49.519
Target Compounds						
21) L5 AR-1248-1	5.659	4.721	16593283	7502927	490.304	478.773
22) L5 AR-1248-2	5.934	4.958	25014694	10795509	483.013	490.556
23) L5 AR-1248-3	6.139	5.000	27267073	11435656	482.490	487.043
24) L5 AR-1248-4	6.543	5.172	27455385	12855546	484.320	481.945
25) L5 AR-1248-5	6.583	5.565	28050962	11642433	474.251	493.393



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092923\
Data File : PO098416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2023 20:12
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 00:46:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.484 min
Delta R.T.: 0.000 min
Response: 88724586
Conc: 46.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: 0.000 min
Response: 36919333
Conc: 47.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.299 min
Delta R.T.: 0.005 min
Response: 58305044
Conc: 48.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.648 min
Delta R.T.: -0.004 min
Response: 26530457
Conc: 49.52 ng/ml



#21 AR-1248-1


R.T.: 5.659 min
Delta R.T.: -0.004 min
Response: 16593283
Conc: 490.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.721 min
Delta R.T.: 0.000 min
Response: 7502927
Conc: 478.77 ng/ml




#22 AR-1248-2

R.T.: 5.934 min
Delta R.T.: -0.004 min
Response: 25014694
Conc: 483.01 ng/ml



#22 AR-1248-2


R.T.: 4.958 min
Delta R.T.: 0.000 min
Response: 10795509
Conc: 490.56 ng/ml



#23 AR-1248-3


R.T.: 6.139 min
Delta R.T.: -0.003 min
Response: 27267073
Conc: 482.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.000 min
Delta R.T.: 0.000 min
Response: 11435656
Conc: 487.04 ng/ml




#24 AR-1248-4

R.T.: 6.543 min
Delta R.T.: -0.005 min
Response: 27455385
Conc: 484.32 ng/ml



#24 AR-1248-4


R.T.: 5.172 min
Delta R.T.: -0.001 min
Response: 12855546
Conc: 481.95 ng/ml



#25 AR-1248-5

R.T.: 6.583 min
Delta R.T.: -0.004 min
Response: 28050962
Conc: 474.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.565 min
Delta R.T.: -0.002 min
Response: 11642433
Conc: 493.39 ng/ml