

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092923\
 Data File : P0098434.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2023 07:08
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 00:56:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.486	3.639	91153212	38131420	47.471	48.958
2) SA Decachlor...	10.300	8.649	60518894	27501608	49.856	51.331
Target Compounds						
16) L4 AR-1242-1	5.661	4.723	23366966	10482777	513.147	500.709
17) L4 AR-1242-2	5.684	4.742	34350486	14446196	501.540	496.827
18) L4 AR-1242-3	5.746	4.918	21682536	8092106	512.030	517.792
19) L4 AR-1242-4	5.844	5.001	16967219	8617980	517.115	520.335
20) L4 AR-1242-5	6.585	5.525	18570291	9672128	524.270	521.170



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092923\
Data File : PO098434.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2023 07:08
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 00:56:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.486 min
Delta R.T.: 0.001 min
Response: 91153212
Conc: 47.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: 0.002 min
Response: 38131420
Conc: 48.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.300 min
Delta R.T.: 0.007 min
Response: 60518894
Conc: 49.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.649 min
Delta R.T.: -0.003 min
Response: 27501608
Conc: 51.33 ng/ml



#16 AR-1242-1


R.T.: 5.661 min
Delta R.T.: 0.001 min
Response: 23366966
Conc: 513.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.723 min
Delta R.T.: 0.000 min
Response: 10482777
Conc: 500.71 ng/ml




#17 AR-1242-2

R.T.: 5.684 min
Delta R.T.: 0.002 min
Response: 34350486
Conc: 501.54 ng/ml



#17 AR-1242-2


R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 14446196
Conc: 496.83 ng/ml



#18 AR-1242-3


R.T.: 5.746 min
Delta R.T.: 0.001 min
Response: 21682536
Conc: 512.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.918 min
Delta R.T.: 0.000 min
Response: 8092106
Conc: 517.79 ng/ml




#19 AR-1242-4

R.T.: 5.844 min
Delta R.T.: 0.000 min
Response: 16967219
Conc: 517.12 ng/ml



#19 AR-1242-4


R.T.: 5.001 min
Delta R.T.: 0.000 min
Response: 8617980
Conc: 520.34 ng/ml



#20 AR-1242-5

R.T.: 6.585 min
Delta R.T.: 0.000 min
Response: 18570291
Conc: 524.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.525 min
Delta R.T.: -0.001 min
Response: 9672128
Conc: 521.17 ng/ml