

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0092925\
 Data File : P0114110.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2025 01:43
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 02:51:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 16 02:49:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.657	3.649	449.1E6	251.3E6	50.915	47.045
2)	SA Decachlor...	8.662	8.607	276.3E6	92351340	41.303	44.235
Target Compounds							
26)	L6 AR-1254-1	5.540	5.513	302.2E6	146.8E6	554.213	485.614
27)	L6 AR-1254-2	5.689	5.661	281.2E6	127.7E6	553.001	483.688
28)	L6 AR-1254-3	6.091	6.060	380.5E6	184.5E6	486.540	476.308
29)	L6 AR-1254-4	6.321	6.287	263.8E6	113.4E6	482.353	477.801
30)	L6 AR-1254-5	6.738	6.702	374.2E6	134.0E6	502.666	436.371



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO092925\
Data File : PO114110.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 01:43
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 02:51:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 16 02:49:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.657 min
Delta R.T.: -0.003 min
Response: 449050872
Conc: 50.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: -0.002 min
Response: 251267382
Conc: 47.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.662 min
Delta R.T.: -0.008 min
Response: 276328502
Conc: 41.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.607 min
Delta R.T.: -0.007 min
Response: 92351340
Conc: 44.23 ng/ml



#26 AR-1254-1


R.T.: 5.540 min
Delta R.T.: -0.006 min
Response: 302167349
Conc: 554.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.513 min
Delta R.T.: -0.006 min
Response: 146848830
Conc: 485.61 ng/ml




#27 AR-1254-2

R.T.: 5.689 min
Delta R.T.: -0.006 min
Response: 281167094
Conc: 553.00 ng/ml



#27 AR-1254-2


R.T.: 5.661 min
Delta R.T.: -0.007 min
Response: 127731796
Conc: 483.69 ng/ml



#28 AR-1254-3


R.T.: 6.091 min
Delta R.T.: -0.007 min
Response: 380525637
Conc: 486.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.060 min
Delta R.T.: -0.006 min
Response: 184482473
Conc: 476.31 ng/ml




#29 AR-1254-4

R.T.: 6.321 min
Delta R.T.: -0.006 min
Response: 263776436
Conc: 482.35 ng/ml



#29 AR-1254-4


R.T.: 6.287 min
Delta R.T.: -0.007 min
Response: 113426107
Conc: 477.80 ng/ml



#30 AR-1254-5

R.T.: 6.738 min
Delta R.T.: -0.006 min
Response: 374152715
Conc: 502.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.702 min
Delta R.T.: -0.006 min
Response: 133953588
Conc: 436.37 ng/ml