

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0093020\
 Data File : P0071808.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Oct 2020 5:35
 Operator : DD\AJ
 Sample : L4161-06 10X
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 KMA696E-END-2-1-2

Manual Integrations
 APPROVED

Ankita
 10/1/2020 5:22:46 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 06:52:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0092320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 23 08:07:44 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.347	3.531	209250	135895	2.438	3.300m#
2)	SA Decachlor...	9.963	8.709	217877	111074	1.896	1.827m
Target Compounds							
26)	L6 AR-1254-1	6.551	5.606	775806	133767	211.855m	56.483 #
27)	L6 AR-1254-2	6.768	5.766	328201	137340	54.434	65.571
28)	L6 AR-1254-3	7.143	6.174	279085	204745	44.080	60.507m#
29)	L6 AR-1254-4	7.437	6.416	281827	121314	45.745m	50.955
30)	L6 AR-1254-5	7.856	6.836	240022	157449	40.385m	48.678

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO093020\
Data File : PO071808.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Oct 2020 5:35
Operator : DD\AJ
Sample : L4161-06 10X
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
KMA696E-END-2-1-2

Manual Integrations
APPROVED

Ankita
10/1/2020 5:22:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 06:52:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092320.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 23 08:07:44 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

