

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100120\
 Data File : PO071846.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Oct 2020 19:04
 Operator : DD\AJ
 Sample : L4198-04
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 RO-21

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 08:42:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 23 08:07:44 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.333	3.516	1794517	949008	20.912	23.042
2) SA Decachlor...	9.944	8.692	2310001	1224320	20.107	20.142

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100120\
Data File : PO071846.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Oct 2020 19:04
Operator : DD\AJ
Sample : L4198-04
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
RO-21

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 08:42:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092320.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 23 08:07:44 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.333 min
Delta R.T.: -0.009 min
Response: 1794517
Conc: 20.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
RO-21




#1 Tetrachloro-m-xylene

R.T.: 3.516 min
Delta R.T.: -0.009 min
Response: 949008
Conc: 23.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.944 min
Delta R.T.: -0.013 min
Response: 2310001
Conc: 20.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.692 min
Delta R.T.: -0.012 min
Response: 1224320
Conc: 20.14 ng/ml