

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100125\
 Data File : PO114166.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Oct 2025 09:57
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 15:26:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 16 02:49:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.656	3.647	440.4E6	245.8E6	49.938	46.013
2) SA Decachlor...	8.663	8.606	356.4E6	106.5E6	53.271	51.000
Target Compounds						
26) L6 AR-1254-1	5.540	5.512	285.2E6	141.5E6	523.136	467.834
27) L6 AR-1254-2	5.690	5.660	260.6E6	122.4E6	512.572	463.367
28) L6 AR-1254-3	6.092	6.059	402.7E6	182.2E6	514.870	470.445
29) L6 AR-1254-4	6.322	6.287	295.4E6	113.4E6	540.260	477.534
30) L6 AR-1254-5	6.739	6.702	378.4E6	137.9E6	508.408	449.104



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100125\
Data File : PO114166.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Oct 2025 09:57
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 15:26:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 16 02:49:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.656 min
Delta R.T.: -0.004 min
Response: 440441413
Conc: 49.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: -0.004 min
Response: 245756060
Conc: 46.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.663 min
Delta R.T.: -0.007 min
Response: 356397944
Conc: 53.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.606 min
Delta R.T.: -0.008 min
Response: 106475488
Conc: 51.00 ng/ml



#26 AR-1254-1


R.T.: 5.540 min
Delta R.T.: -0.006 min
Response: 285223599
Conc: 523.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.512 min
Delta R.T.: -0.006 min
Response: 141472002
Conc: 467.83 ng/ml




#27 AR-1254-2

R.T.: 5.690 min
Delta R.T.: -0.006 min
Response: 260611527
Conc: 512.57 ng/ml



#27 AR-1254-2


R.T.: 5.660 min
Delta R.T.: -0.007 min
Response: 122365563
Conc: 463.37 ng/ml



#28 AR-1254-3


R.T.: 6.092 min
Delta R.T.: -0.006 min
Response: 402682872
Conc: 514.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.059 min
Delta R.T.: -0.007 min
Response: 182211757
Conc: 470.45 ng/ml




#29 AR-1254-4

R.T.: 6.322 min
Delta R.T.: -0.006 min
Response: 295443428
Conc: 540.26 ng/ml



#29 AR-1254-4


R.T.: 6.287 min
Delta R.T.: -0.007 min
Response: 113362946
Conc: 477.53 ng/ml



#30 AR-1254-5

R.T.: 6.739 min
Delta R.T.: -0.006 min
Response: 378427016
Conc: 508.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.702 min
Delta R.T.: -0.006 min
Response: 137862434
Conc: 449.10 ng/ml