

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100125\
 Data File : PO114172.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Oct 2025 13:23
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 15:33:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 16 02:49:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.656	3.647	437.2E6	246.8E6	49.576	46.213
2) SA Decachlor...	8.662	8.606	362.6E6	106.8E6	54.193	51.162
Target Compounds						
26) L6 AR-1254-1	5.540	5.512	285.0E6	142.6E6	522.676	471.583
27) L6 AR-1254-2	5.689	5.660	263.0E6	122.7E6	517.180	464.733
28) L6 AR-1254-3	6.092	6.059	402.8E6	182.7E6	515.051	471.811
29) L6 AR-1254-4	6.321	6.287	294.8E6	113.4E6	538.994	477.857
30) L6 AR-1254-5	6.739	6.701	379.1E6	139.1E6	509.332	453.017



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100125\
Data File : PO114172.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Oct 2025 13:23
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 15:33:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091525.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 16 02:49:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.656 min
Delta R.T.: -0.004 min
Response: 437241619
Conc: 49.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: -0.004 min
Response: 246824601
Conc: 46.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.662 min
Delta R.T.: -0.008 min
Response: 362565133
Conc: 54.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.606 min
Delta R.T.: -0.008 min
Response: 106814069
Conc: 51.16 ng/ml



#26 AR-1254-1


R.T.: 5.540 min
Delta R.T.: -0.005 min
Response: 284972750
Conc: 522.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.512 min
Delta R.T.: -0.007 min
Response: 142605674
Conc: 471.58 ng/ml




#27 AR-1254-2

R.T.: 5.689 min
Delta R.T.: -0.006 min
Response: 262954224
Conc: 517.18 ng/ml



#27 AR-1254-2


R.T.: 5.660 min
Delta R.T.: -0.007 min
Response: 122726174
Conc: 464.73 ng/ml



#28 AR-1254-3


R.T.: 6.092 min
Delta R.T.: -0.006 min
Response: 402823943
Conc: 515.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.059 min
Delta R.T.: -0.007 min
Response: 182740555
Conc: 471.81 ng/ml




#29 AR-1254-4

R.T.: 6.321 min
Delta R.T.: -0.006 min
Response: 294750637
Conc: 538.99 ng/ml



#29 AR-1254-4


R.T.: 6.287 min
Delta R.T.: -0.007 min
Response: 113439604
Conc: 477.86 ng/ml



#30 AR-1254-5

R.T.: 6.739 min
Delta R.T.: -0.006 min
Response: 379114565
Conc: 509.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.701 min
Delta R.T.: -0.007 min
Response: 139063473
Conc: 453.02 ng/ml