

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100223\
 Data File : PO098454.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2023 14:10
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 14:24:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.486	3.639	90882871	36817882	47.330	47.271
2)	SA Decachlor...	10.301	8.650	61650243	27536451	50.788	51.396
Target Compounds							
26)	L6 AR-1254-1	6.519	5.525	30234688	17954452	478.672	468.247
27)	L6 AR-1254-2	6.739	5.673	45101072	15827040	469.962	459.717
28)	L6 AR-1254-3	7.106	6.077	45188410	24081279	480.255	463.442
29)	L6 AR-1254-4	7.393	6.306	29465617	13275580	506.070	483.657
30)	L6 AR-1254-5	7.813	6.725	34638704	20855403	495.938	481.449



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100223\
Data File : PO098454.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2023 14:10
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 14:24:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.486 min
Delta R.T.: 0.001 min
Response: 90882871
Conc: 47.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: 0.000 min
Response: 36817882
Conc: 47.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.301 min
Delta R.T.: 0.008 min
Response: 61650243
Conc: 50.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.650 min
Delta R.T.: -0.002 min
Response: 27536451
Conc: 51.40 ng/ml



#26 AR-1254-1


R.T.: 6.519 min
Delta R.T.: -0.003 min
Response: 30234688
Conc: 478.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.525 min
Delta R.T.: 0.000 min
Response: 17954452
Conc: 468.25 ng/ml




#27 AR-1254-2

R.T.: 6.739 min
Delta R.T.: -0.003 min
Response: 45101072
Conc: 469.96 ng/ml



#27 AR-1254-2


R.T.: 5.673 min
Delta R.T.: 0.000 min
Response: 15827040
Conc: 459.72 ng/ml



#28 AR-1254-3


R.T.: 7.106 min
Delta R.T.: -0.003 min
Response: 45188410
Conc: 480.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.077 min
Delta R.T.: -0.001 min
Response: 24081279
Conc: 463.44 ng/ml




#29 AR-1254-4

R.T.: 7.393 min
Delta R.T.: -0.003 min
Response: 29465617
Conc: 506.07 ng/ml



#29 AR-1254-4


R.T.: 6.306 min
Delta R.T.: 0.000 min
Response: 13275580
Conc: 483.66 ng/ml



#30 AR-1254-5

R.T.: 7.813 min
Delta R.T.: -0.002 min
Response: 34638704
Conc: 495.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.725 min
Delta R.T.: 0.000 min
Response: 20855403
Conc: 481.45 ng/ml