

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100320\
 Data File : PO071931.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Oct 2020 21:07
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 04:20:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 03 04:09:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.511	1597287	814780	22.238	21.753
2) SA Decachlor...	9.936	8.685	2400597	1264387	21.802	22.229
Target Compounds						
9) L2 AR-1221-2	0.000	3.850	0	11889	N.D.	36.809 #
29) L6 AR-1254-4	7.424	0.000	40604	0	7.507	N.D. #
30) L6 AR-1254-5	0.000	6.819	0	10744	N.D.	3.363 #
36) L8 AR-1262-1	0.000	6.819	0	10744	N.D.	6.066 #
45) L9 AR-1268-5	0.000	8.471	0	11083	N.D.	0.586 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100320\
Data File : PO071931.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Oct 2020 21:07
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 04:20:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100320.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 03 04:09:16 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: 0.000 min
Response: 1597287
Conc: 22.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.511 min
Delta R.T.: 0.000 min
Response: 814780
Conc: 21.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.936 min
Delta R.T.: 0.000 min
Response: 2400597
Conc: 21.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 1264387
Conc: 22.23 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 4.681 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.850 min
Delta R.T.: -0.009 min
Response: 11889
Conc: 36.81 ng/ml




#29 AR-1254-4

R.T.: 7.424 min
Delta R.T.: 0.002 min
Response: 40604
Conc: 7.51 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.397 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 7.842 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#30 AR-1254-5

R.T.: 6.819 min
Delta R.T.: 0.000 min
Response: 10744
Conc: 3.36 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 7.911 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.819 min
Delta R.T.: 0.002 min
Response: 10744
Conc: 6.07 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.676 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.471 min
Delta R.T.: -0.003 min
Response: 11083
Conc: 0.59 ng/ml