

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100323\
 Data File : PO098500.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Oct 2023 00:04
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/04/2023
 Supervised By :Ankita Jodhani 10/04/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 01:37:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.483	3.635	91919027	38447355	47.870	49.364
2) SA Decachlor...	10.295	8.640	60403029	27677954	49.761	51.661m
Target Compounds						
3) L1 AR-1016-1	5.658	4.718	26189969	11889156	484.607	498.116
4) L1 AR-1016-2	5.679	4.736	40184557	16805658	494.957	493.548
5) L1 AR-1016-3	5.742	4.912	25131810	9181210	502.305	510.390
6) L1 AR-1016-4	5.841	4.954	19556869	7682733	500.155	512.479
7) L1 AR-1016-5	6.137	5.168	20460346	9311073	497.565	479.448
31) L7 AR-1260-1	7.267	6.201	36127739	18272987	496.841	498.030
32) L7 AR-1260-2	7.524	6.390	40737147	21187685	494.738	504.944
33) L7 AR-1260-3	7.884	6.544	32026979	20074115	517.712	501.639
34) L7 AR-1260-4	8.112	7.015	35984097	16282333	533.789	505.614
35) L7 AR-1260-5	8.437	7.257	63271696	34379072	487.006	490.706

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100323\
Data File : PO098500.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2023 00:04
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/04/2023
Supervised By :Ankita Jodhani 10/04/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 01:37:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

