

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100421\
 Data File : PO081727.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Oct 2021 11:59
 Operator : AJ\MA
 Sample : M4040-07
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SW-RR-50

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 16:05:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 02:21:54 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.868	3.984	1464609	415621	16.424	15.600
2)	SA Decachlor...	10.876	9.328	1716935	538332	23.050	23.181
Target Compounds							
3)	L1 AR-1016-1	6.224f	5.244	-27463	19818	N.D.	21.997 #
4)	L1 AR-1016-2	6.224	5.304f	-27463	58409	N.D.	42.772 #
5)	L1 AR-1016-3	6.265f	5.489	193446	47974	67.568	67.666
6)	L1 AR-1016-4	6.383	5.543f	175141	65883	76.140	106.362 #
7)	L1 AR-1016-5	6.702	5.747	830903	-23159	365.120	N.D. #
8)	L2 AR-1221-1	5.116	0.000	3032068	0	2770.803	N.D. #
9)	L2 AR-1221-2	5.235	0.000	918575	0	1165.029	N.D. #
10)	L2 AR-1221-3	5.330f	4.424	991953	972421	394.258	1253.239 #
11)	L3 AR-1232-1	5.330f	4.424	991953	972421	508.951	1621.801 #
12)	L3 AR-1232-2	5.922f	5.304f	503945	58409	473.739	92.302 #
13)	L3 AR-1232-3	6.224	5.489f	-27463	47974	N.D.	151.084 #
14)	L3 AR-1232-4	6.383	5.574	175141	39743	171.394	132.258
15)	L3 AR-1232-5	6.484	5.747	1225101	-23159	1661.531	N.D. #
16)	L4 AR-1242-1	6.224f	5.244	-27463	19818	N.D.	27.014 #
17)	L4 AR-1242-2	6.224	5.304f	-27463	58409	N.D.	53.049 #
18)	L4 AR-1242-3	6.265f	5.489f	193446	47974	84.771	85.576
19)	L4 AR-1242-4	6.383	5.574	175141	39743	96.441	69.215 #
20)	L4 AR-1242-5	7.173	6.105f	459074	350591	232.793	469.160 #
21)	L5 AR-1248-1	6.224f	5.244	-27463	19818	N.D.	33.114 #
22)	L5 AR-1248-2	6.484	5.543	1225101	65883	464.488	74.245 #
23)	L5 AR-1248-3	6.702	5.574	830903	39743	265.478	45.459 #
24)	L5 AR-1248-4	7.140	5.747	246723	-23159	71.635	N.D. #
25)	L5 AR-1248-5	7.173	6.191	459074	90843	135.788	94.130 #
26)	L6 AR-1254-1	7.092f	0.000	275058	0	76.715	N.D. #
27)	L6 AR-1254-2	7.333	6.269f	90815	41548	16.470	30.538 #
28)	L6 AR-1254-3	7.737	6.718	878593	208554	155.709	104.845 #
29)	L6 AR-1254-4	8.016	6.968	618814	190227	150.487	152.348
30)	L6 AR-1254-5	8.445	7.386	2034028	91225	441.280	48.799 #
31)	L7 AR-1260-1	7.871f	0.000	1230256	0	291.030	N.D. #
32)	L7 AR-1260-2	8.179f	7.072f	3076361	263146	592.697	161.012 #
33)	L7 AR-1260-3	8.530	7.201	872694	798162	269.331	479.984 #
34)	L7 AR-1260-4	8.736f	7.677f	697903	192577	175.087	173.266
35)	L7 AR-1260-5	9.102f	7.945	623277	378964	84.134	153.451 #
36)	L8 AR-1262-1	8.530	7.386	872694	91225	168.364	91.651 #
37)	L8 AR-1262-2	9.102f	7.945	623277	378964	69.882	123.697 #
38)	L8 AR-1262-3	0.000	8.237	0	81049	N.D.	60.560 #
39)	L8 AR-1262-4	0.000	8.291	0	6567	N.D.	2.791 #
40)	L8 AR-1262-5	10.117f	0.000	184375	0	54.109	N.D. #
41)	L9 AR-1268-1	0.000	8.237	0	81049	N.D.	21.135 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100421\
 Data File : P0081727.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Oct 2021 11:59
 Operator : AJ\MA
 Sample : M4040-07
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SW-RR-50

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 16:05:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 02:21:54 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	0.000	8.291	0	6567	N.D.	1.886 #
43) L9	AR-1268-3	9.691	8.510	96688	35733	11.285	11.381
44) L9	AR-1268-4	10.117f	0.000	184375	0	48.018	N.D. #
45) L9	AR-1268-5	0.000	9.088	0	88896	N.D.	9.888 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100421\
Data File : PO081727.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Oct 2021 11:59
Operator : AJ\MA
Sample : M4040-07
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
SW-RR-50

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 16:05:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 24 02:21:54 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.868 min
Delta R.T.: -0.003 min
Response: 1464609
Conc: 16.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#1 Tetrachloro-m-xylene

R.T.: 3.984 min
Delta R.T.: -0.005 min
Response: 415621
Conc: 15.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.876 min
Delta R.T.: -0.010 min
Response: 1716935
Conc: 23.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.328 min
Delta R.T.: -0.009 min
Response: 538332
Conc: 23.18 ng/ml



#3 AR-1016-1


R.T.: 6.224 min
Delta R.T.: 0.025 min
Response: -27463
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#3 AR-1016-1

R.T.: 5.244 min
Delta R.T.: -0.013 min
Response: 19818
Conc: 22.00 ng/ml




#4 AR-1016-2

R.T.: 6.224 min
Delta R.T.: 0.001 min
Response: -27463
Conc: N.D.



#4 AR-1016-2


R.T.: 5.304 min
Delta R.T.: 0.027 min
Response: 58409
Conc: 42.77 ng/ml



#5 AR-1016-3


R.T.: 6.265 min
Delta R.T.: -0.022 min
Response: 193446
Conc: 67.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#5 AR-1016-3

R.T.: 5.489 min
Delta R.T.: 0.019 min
Response: 47974
Conc: 67.67 ng/ml




#6 AR-1016-4

R.T.: 6.383 min
Delta R.T.: -0.012 min
Response: 175141
Conc: 76.14 ng/ml



#6 AR-1016-4


R.T.: 5.543 min
Delta R.T.: 0.020 min
Response: 65883
Conc: 106.36 ng/ml



#7 AR-1016-5


R.T.: 6.702 min
Delta R.T.: -0.009 min
Response: 830903
Conc: 365.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#7 AR-1016-5

R.T.: 5.747 min
Delta R.T.: -0.008 min
Response: -23159
Conc: N.D.




#8 AR-1221-1

R.T.: 5.116 min
Delta R.T.: -0.007 min
Response: 3032068
Conc: 2770.80 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.248 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 5.235 min
Delta R.T.: 0.013 min
Response: 918575
Conc: 1165.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.348 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 5.330 min
Delta R.T.: 0.020 min
Response: 991953
Conc: 394.26 ng/ml



#10 AR-1221-3


R.T.: 4.424 min
Delta R.T.: -0.013 min
Response: 972421
Conc: 1253.24 ng/ml



#11 AR-1232-1


R.T.: 5.330 min
Delta R.T.: 0.021 min
Response: 991953
Conc: 508.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#11 AR-1232-1

R.T.: 4.424 min
Delta R.T.: -0.012 min
Response: 972421
Conc: 1621.80 ng/ml




#12 AR-1232-2

R.T.: 5.922 min
Delta R.T.: 0.022 min
Response: 503945
Conc: 473.74 ng/ml



#12 AR-1232-2


R.T.: 5.304 min
Delta R.T.: 0.029 min
Response: 58409
Conc: 92.30 ng/ml



#13 AR-1232-3


R.T.: 6.224 min
Delta R.T.: 0.004 min
Response: -27463
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#13 AR-1232-3

R.T.: 5.489 min
Delta R.T.: 0.020 min
Response: 47974
Conc: 151.08 ng/ml




#14 AR-1232-4

R.T.: 6.383 min
Delta R.T.: -0.011 min
Response: 175141
Conc: 171.39 ng/ml



#14 AR-1232-4


R.T.: 5.574 min
Delta R.T.: 0.008 min
Response: 39743
Conc: 132.26 ng/ml



#15 AR-1232-5


R.T.: 6.484 min
Delta R.T.: -0.008 min
Response: 1225101
Conc: 1661.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#15 AR-1232-5

R.T.: 5.747 min
Delta R.T.: -0.006 min
Response: -23159
Conc: N.D.




#16 AR-1242-1

R.T.: 6.224 min
Delta R.T.: 0.028 min
Response: -27463
Conc: N.D.



#16 AR-1242-1


R.T.: 5.244 min
Delta R.T.: -0.012 min
Response: 19818
Conc: 27.01 ng/ml



#17 AR-1242-2


R.T.: 6.224 min
Delta R.T.: 0.004 min
Response: -27463
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#17 AR-1242-2

R.T.: 5.304 min
Delta R.T.: 0.028 min
Response: 58409
Conc: 53.05 ng/ml




#18 AR-1242-3

R.T.: 6.265 min
Delta R.T.: -0.021 min
Response: 193446
Conc: 84.77 ng/ml



#18 AR-1242-3


R.T.: 5.489 min
Delta R.T.: 0.020 min
Response: 47974
Conc: 85.58 ng/ml



#19 AR-1242-4


R.T.: 6.383 min
Delta R.T.: -0.010 min
Response: 175141
Conc: 96.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#19 AR-1242-4

R.T.: 5.574 min
Delta R.T.: 0.008 min
Response: 39743
Conc: 69.21 ng/ml




#20 AR-1242-5

R.T.: 7.173 min
Delta R.T.: -0.006 min
Response: 459074
Conc: 232.79 ng/ml



#20 AR-1242-5


R.T.: 6.105 min
Delta R.T.: -0.030 min
Response: 350591
Conc: 469.16 ng/ml



#21 AR-1248-1


R.T.: 6.224 min
Delta R.T.: 0.027 min
Response: -27463
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#21 AR-1248-1

R.T.: 5.244 min
Delta R.T.: -0.012 min
Response: 19818
Conc: 33.11 ng/ml




#22 AR-1248-2

R.T.: 6.484 min
Delta R.T.: -0.009 min
Response: 1225101
Conc: 464.49 ng/ml



#22 AR-1248-2


R.T.: 5.543 min
Delta R.T.: 0.021 min
Response: 65883
Conc: 74.24 ng/ml



#23 AR-1248-3


R.T.: 6.702 min
Delta R.T.: -0.009 min
Response: 830903
Conc: 265.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#23 AR-1248-3

R.T.: 5.574 min
Delta R.T.: 0.008 min
Response: 39743
Conc: 45.46 ng/ml




#24 AR-1248-4

R.T.: 7.140 min
Delta R.T.: 0.000 min
Response: 246723
Conc: 71.63 ng/ml



#24 AR-1248-4


R.T.: 5.747 min
Delta R.T.: -0.006 min
Response: -23159
Conc: N.D.



#25 AR-1248-5


R.T.: 7.173 min
Delta R.T.: -0.008 min
Response: 459074
Conc: 135.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#25 AR-1248-5

R.T.: 6.191 min
Delta R.T.: 0.012 min
Response: 90843
Conc: 94.13 ng/ml




#26 AR-1254-1

R.T.: 7.092 min
Delta R.T.: -0.018 min
Response: 275058
Conc: 76.71 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T.: 6.136 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 7.333 min
Delta R.T.: -0.008 min
Response: 90815
Conc: 16.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#27 AR-1254-2

R.T.: 6.269 min
Delta R.T.: -0.028 min
Response: 41548
Conc: 30.54 ng/ml




#28 AR-1254-3

R.T.: 7.737 min
Delta R.T.: 0.013 min
Response: 878593
Conc: 155.71 ng/ml



#28 AR-1254-3


R.T.: 6.718 min
Delta R.T.: -0.003 min
Response: 208554
Conc: 104.84 ng/ml



#29 AR-1254-4


R.T.: 8.016 min
Delta R.T.: -0.005 min
Response: 618814
Conc: 150.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#29 AR-1254-4

R.T.: 6.968 min
Delta R.T.: 0.006 min
Response: 190227
Conc: 152.35 ng/ml




#30 AR-1254-5

R.T.: 8.445 min
Delta R.T.: -0.005 min
Response: 2034028
Conc: 441.28 ng/ml



#30 AR-1254-5


R.T.: 7.386 min
Delta R.T.: -0.008 min
Response: 91225
Conc: 48.80 ng/ml



#31 AR-1260-1


R.T.: 7.871 min
Delta R.T.: -0.019 min
Response: 1230256
Conc: 291.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.858 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 8.179 min
Delta R.T.: 0.021 min
Response: 3076361
Conc: 592.70 ng/ml



#32 AR-1260-2


R.T.: 7.072 min
Delta R.T.: 0.015 min
Response: 263146
Conc: 161.01 ng/ml



#33 AR-1260-3


R.T.: 8.530 min
Delta R.T.: 0.007 min
Response: 872694
Conc: 269.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#33 AR-1260-3

R.T.: 7.201 min
Delta R.T.: -0.012 min
Response: 798162
Conc: 479.98 ng/ml




#34 AR-1260-4

R.T.: 8.736 min
Delta R.T.: -0.018 min
Response: 697903
Conc: 175.09 ng/ml



#34 AR-1260-4


R.T.: 7.677 min
Delta R.T.: -0.023 min
Response: 192577
Conc: 173.27 ng/ml



#35 AR-1260-5


R.T.: 9.102 min
Delta R.T.: 0.024 min
Response: 623277
Conc: 84.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#35 AR-1260-5

R.T.: 7.945 min
Delta R.T.: -0.005 min
Response: 378964
Conc: 153.45 ng/ml




#36 AR-1262-1

R.T.: 8.530 min
Delta R.T.: 0.007 min
Response: 872694
Conc: 168.36 ng/ml



#36 AR-1262-1

R.T.: 7.386 min
Delta R.T.: -0.007 min
Response: 91225
Conc: 91.65 ng/ml



#37 AR-1262-2

R.T.: 9.102 min
Delta R.T.: 0.023 min
Response: 623277
Conc: 69.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50

#37 AR-1262-2


R.T.: 7.945 min
Delta R.T.: -0.004 min
Response: 378964
Conc: 123.70 ng/ml

#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 9.385 min
Response: 0
Conc: N.D.

#38 AR-1262-3


R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 81049
Conc: 60.56 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 9.480 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#39 AR-1262-4

R.T.: 8.291 min
Delta R.T.: -0.011 min
Response: 6567
Conc: 2.79 ng/ml




#40 AR-1262-5

R.T.: 10.117 min
Delta R.T.: -0.024 min
Response: 184375
Conc: 54.11 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 8.804 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 9.383 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#41 AR-1268-1

R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 81049
Conc: 21.13 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 9.480 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 8.291 min
Delta R.T.: -0.011 min
Response: 6567
Conc: 1.89 ng/ml



#43 AR-1268-3


R.T.: 9.691 min
Delta R.T.: -0.012 min
Response: 96688
Conc: 11.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
SW-RR-50




#43 AR-1268-3

R.T.: 8.510 min
Delta R.T.: 0.000 min
Response: 35733
Conc: 11.38 ng/ml




#44 AR-1268-4

R.T.: 10.117 min
Delta R.T.: -0.023 min
Response: 184375
Conc: 48.02 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 8.803 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.550 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SW-RR-50



#45 AR-1268-5

R.T.: 9.088 min
Delta R.T.: 0.000 min
Response: 88896
Conc: 9.89 ng/ml