

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100421\
 Data File : P0081748.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Oct 2021 18:37
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 01:06:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 02:21:54 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.869	3.985	4540391	1274357	50.915	47.833
2) SA Decachlor...	10.878	9.328	3742022	1135409	50.237	48.892
Target Compounds						
21) L5 AR-1248-1	6.197	5.252	1018821	280721	506.151	469.068
22) L5 AR-1248-2	6.491	5.517	1326222	403749	502.828	454.990
23) L5 AR-1248-3	6.709	5.561	1572659	401356	502.472	459.074
24) L5 AR-1248-4	7.140	5.748	1709955	467590	496.476	460.200
25) L5 AR-1248-5	7.178	6.174	1627121	442365	481.279	458.367



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100421\
Data File : PO081748.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Oct 2021 18:37
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 01:06:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 24 02:21:54 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.869 min
Delta R.T.: -0.002 min
Response: 4540391
Conc: 50.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.985 min
Delta R.T.: -0.004 min
Response: 1274357
Conc: 47.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.878 min
Delta R.T.: -0.008 min
Response: 3742022
Conc: 50.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.328 min
Delta R.T.: -0.009 min
Response: 1135409
Conc: 48.89 ng/ml



#21 AR-1248-1


R.T.: 6.197 min
Delta R.T.: 0.000 min
Response: 1018821
Conc: 506.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.252 min
Delta R.T.: -0.005 min
Response: 280721
Conc: 469.07 ng/ml




#22 AR-1248-2

R.T.: 6.491 min
Delta R.T.: -0.002 min
Response: 1326222
Conc: 502.83 ng/ml



#22 AR-1248-2


R.T.: 5.517 min
Delta R.T.: -0.005 min
Response: 403749
Conc: 454.99 ng/ml



#23 AR-1248-3


R.T.: 6.709 min
Delta R.T.: -0.002 min
Response: 1572659
Conc: 502.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.561 min
Delta R.T.: -0.005 min
Response: 401356
Conc: 459.07 ng/ml




#24 AR-1248-4

R.T.: 7.140 min
Delta R.T.: 0.000 min
Response: 1709955
Conc: 496.48 ng/ml



#24 AR-1248-4


R.T.: 5.748 min
Delta R.T.: -0.005 min
Response: 467590
Conc: 460.20 ng/ml



#25 AR-1248-5

R.T.: 7.178 min
Delta R.T.: -0.002 min
Response: 1627121
Conc: 481.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.174 min
Delta R.T.: -0.004 min
Response: 442365
Conc: 458.37 ng/ml