

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO100423\
 Data File : PO098546.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Oct 2023 21:24
 Operator : YP/AJ
 Sample : 04725-08
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 K242-3A

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/05/2023
 Supervised By :Ankita Jodhani 10/05/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 00:14:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.481	3.634	35640368	15156879	18.561	19.460
2) SA Decachlor...	10.293	8.639	17596503	8669527	14.496	16.182m
Target Compounds						
26) L6 AR-1254-1	6.507	5.520	6082337	1954542	96.295	50.974m#
27) L6 AR-1254-2	6.735	5.666	4466726	1176850	46.544	34.183m#
28) L6 AR-1254-3	7.103	6.070	5337075	2446827	56.722	47.089m
29) L6 AR-1254-4	7.389	6.300	2407592	1290380	41.350	47.011
30) L6 AR-1254-5	7.814	6.717	5619858	8007761	80.462m	184.860 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100423\
Data File : PO098546.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2023 21:24
Operator : YP/AJ
Sample : 04725-08
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
K242-3A

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/05/2023
Supervised By :Ankita Jodhani 10/05/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 00:14:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

