

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100423\
 Data File : PO098542.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Oct 2023 20:16
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 00:12:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.481	3.634	88715375	38735880	46.202	49.734
2) SA Decachlor...	10.295	8.640	58703559	27998993	48.361	52.260m
Target Compounds						
21) L5 AR-1248-1	5.655	4.717	16623585	7810228	491.200	498.382
22) L5 AR-1248-2	5.930	4.953	24755553	11103065	478.009	504.531
23) L5 AR-1248-3	6.136	4.995	26906057	12408569	476.102	528.479
24) L5 AR-1248-4	6.541	5.167	27990153	14024908	493.754	525.784
25) L5 AR-1248-5	6.580	5.560	29307001	12550125	495.487	531.860



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100423\
Data File : PO098542.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2023 20:16
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 00:12:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.481 min
Delta R.T.: -0.004 min
Response: 88715375
Conc: 46.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.634 min
Delta R.T.: -0.004 min
Response: 38735880
Conc: 49.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.295 min
Delta R.T.: 0.001 min
Response: 58703559
Conc: 48.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.640 min
Delta R.T.: -0.012 min
Response: 27998993
Conc: 52.26 ng/ml m



#21 AR-1248-1


R.T.: 5.655 min
Delta R.T.: -0.007 min
Response: 16623585
Conc: 491.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.717 min
Delta R.T.: -0.005 min
Response: 7810228
Conc: 498.38 ng/ml




#22 AR-1248-2

R.T.: 5.930 min
Delta R.T.: -0.007 min
Response: 24755553
Conc: 478.01 ng/ml



#22 AR-1248-2


R.T.: 4.953 min
Delta R.T.: -0.005 min
Response: 11103065
Conc: 504.53 ng/ml



#23 AR-1248-3


R.T.: 6.136 min
Delta R.T.: -0.007 min
Response: 26906057
Conc: 476.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.995 min
Delta R.T.: -0.006 min
Response: 12408569
Conc: 528.48 ng/ml




#24 AR-1248-4

R.T.: 6.541 min
Delta R.T.: -0.008 min
Response: 27990153
Conc: 493.75 ng/ml



#24 AR-1248-4


R.T.: 5.167 min
Delta R.T.: -0.006 min
Response: 14024908
Conc: 525.78 ng/ml



#25 AR-1248-5

R.T.: 6.580 min
Delta R.T.: -0.007 min
Response: 29307001
Conc: 495.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.560 min
Delta R.T.: -0.007 min
Response: 12550125
Conc: 531.86 ng/ml