

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
 Data File : PO089758.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2022 01:42
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 03:41:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 05 06:04:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.335	3.561	150.1E6	53073232	57.752	58.768
2) SA Decachlor...	10.094	8.591	65412292	43082062	60.851	54.217
Target Compounds						
16) L4 AR-1242-1	5.504	4.652	34432992	14971009	559.428	554.685
17) L4 AR-1242-2	5.526	4.671	48231815	19852953	545.042	551.330
18) L4 AR-1242-3	5.589	4.847	30733761	10867490	554.658	565.006
19) L4 AR-1242-4	5.687	4.932	24425559	10946734	567.503	565.222
20) L4 AR-1242-5	6.428	5.457	24042290	13468075	581.727	558.891

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
Data File : PO089758.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2022 01:42
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 03:41:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 05 06:04:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

