

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
 Data File : PO089730.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 16:15
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 18:19:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 05 06:04:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.336	3.561	134.0E6	51038690	51.525	56.515
2) SA Decachlor...	10.098	8.593	58774017	39016233	54.675	49.101
Target Compounds						
21) L5 AR-1248-1	5.506	4.652	23481856	10694455	516.224	547.491
22) L5 AR-1248-2	5.781	4.891	32217372	13895086	498.611	512.843
23) L5 AR-1248-3	5.986	4.933	34210009	14994229	502.340	525.452
24) L5 AR-1248-4	6.391	5.105	36627244	17073522	507.249	508.327
25) L5 AR-1248-5	6.430	5.500	35978734	17083115	515.894	524.436



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
Data File : PO089730.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 16:15
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 18:19:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 05 06:04:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.336 min
Delta R.T.: 0.001 min
Response: 133954811
Conc: 51.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.561 min
Delta R.T.: 0.005 min
Response: 51038690
Conc: 56.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.098 min
Delta R.T.: 0.006 min
Response: 58774017
Conc: 54.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.593 min
Delta R.T.: 0.005 min
Response: 39016233
Conc: 49.10 ng/ml



#21 AR-1248-1


R.T.: 5.506 min
Delta R.T.: 0.001 min
Response: 23481856
Conc: 516.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.652 min
Delta R.T.: 0.003 min
Response: 10694455
Conc: 547.49 ng/ml




#22 AR-1248-2

R.T.: 5.781 min
Delta R.T.: 0.002 min
Response: 32217372
Conc: 498.61 ng/ml



#22 AR-1248-2


R.T.: 4.891 min
Delta R.T.: 0.003 min
Response: 13895086
Conc: 512.84 ng/ml



#23 AR-1248-3


R.T.: 5.986 min
Delta R.T.: 0.001 min
Response: 34210009
Conc: 502.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.933 min
Delta R.T.: 0.003 min
Response: 14994229
Conc: 525.45 ng/ml




#24 AR-1248-4

R.T.: 6.391 min
Delta R.T.: 0.001 min
Response: 36627244
Conc: 507.25 ng/ml



#24 AR-1248-4


R.T.: 5.105 min
Delta R.T.: 0.004 min
Response: 17073522
Conc: 508.33 ng/ml



#25 AR-1248-5

R.T.: 6.430 min
Delta R.T.: 0.002 min
Response: 35978734
Conc: 515.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.500 min
Delta R.T.: 0.003 min
Response: 17083115
Conc: 524.44 ng/ml