

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
 Data File : PO089731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 16:32
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 18:20:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 05 06:04:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.335	3.561	146.2E6	54611428	56.244	60.472
2) SA Decachlor...	10.096	8.594	62652130	42474080	58.283	53.452
Target Compounds						
26) L6 AR-1254-1	6.364	5.459	38122937	27454457	522.538	545.805
27) L6 AR-1254-2	6.585	5.608	56495618	23620426	531.311	540.522
28) L6 AR-1254-3	6.952	6.014	56392202	38163135	517.613	537.153
29) L6 AR-1254-4	7.238	6.243	39349725	24017209	528.982	542.333
30) L6 AR-1254-5	7.660	6.663	40565515	32675065	534.584	538.780



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
Data File : PO089731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 16:32
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 18:20:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 05 06:04:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.335 min
Delta R.T.: 0.000 min
Response: 146225744
Conc: 56.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.561 min
Delta R.T.: 0.005 min
Response: 54611428
Conc: 60.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.096 min
Delta R.T.: 0.004 min
Response: 62652130
Conc: 58.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.594 min
Delta R.T.: 0.006 min
Response: 42474080
Conc: 53.45 ng/ml



#26 AR-1254-1


R.T.: 6.364 min
Delta R.T.: 0.000 min
Response: 38122937
Conc: 522.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.459 min
Delta R.T.: 0.004 min
Response: 27454457
Conc: 545.80 ng/ml




#27 AR-1254-2

R.T.: 6.585 min
Delta R.T.: 0.001 min
Response: 56495618
Conc: 531.31 ng/ml



#27 AR-1254-2


R.T.: 5.608 min
Delta R.T.: 0.004 min
Response: 23620426
Conc: 540.52 ng/ml



#28 AR-1254-3


R.T.: 6.952 min
Delta R.T.: 0.002 min
Response: 56392202
Conc: 517.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.014 min
Delta R.T.: 0.004 min
Response: 38163135
Conc: 537.15 ng/ml




#29 AR-1254-4

R.T.: 7.238 min
Delta R.T.: 0.001 min
Response: 39349725
Conc: 528.98 ng/ml



#29 AR-1254-4


R.T.: 6.243 min
Delta R.T.: 0.004 min
Response: 24017209
Conc: 542.33 ng/ml



#30 AR-1254-5

R.T.: 7.660 min
Delta R.T.: 0.001 min
Response: 40565515
Conc: 534.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.663 min
Delta R.T.: 0.003 min
Response: 32675065
Conc: 538.78 ng/ml