

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
 Data File : PO089745.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 21:06
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 23:10:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 05 06:04:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.335	3.561	131.2E6	47971880	50.461	53.120
2) SA Decachlor...	10.094	8.593	57283040	38718075	53.288	48.725
Target Compounds						
26) L6 AR-1254-1	6.364	5.459	35189805	24608241	482.334	489.221
27) L6 AR-1254-2	6.585	5.607	52209680	21351679	491.004	488.604
28) L6 AR-1254-3	6.951	6.013	52461824	34376057	481.537	483.850
29) L6 AR-1254-4	7.238	6.243	36730376	21471276	493.770	484.843
30) L6 AR-1254-5	7.660	6.663	37552734	29319445	494.881	483.449



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
Data File : PO089745.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 21:06
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 23:10:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 05 06:04:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.335 min
Delta R.T.: 0.000 min
Response: 131191050
Conc: 50.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.561 min
Delta R.T.: 0.005 min
Response: 47971880
Conc: 53.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.094 min
Delta R.T.: 0.002 min
Response: 57283040
Conc: 53.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.593 min
Delta R.T.: 0.005 min
Response: 38718075
Conc: 48.73 ng/ml



#26 AR-1254-1


R.T.: 6.364 min
Delta R.T.: 0.000 min
Response: 35189805
Conc: 482.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.459 min
Delta R.T.: 0.004 min
Response: 24608241
Conc: 489.22 ng/ml




#27 AR-1254-2

R.T.: 6.585 min
Delta R.T.: 0.001 min
Response: 52209680
Conc: 491.00 ng/ml



#27 AR-1254-2


R.T.: 5.607 min
Delta R.T.: 0.003 min
Response: 21351679
Conc: 488.60 ng/ml



#28 AR-1254-3


R.T.: 6.951 min
Delta R.T.: 0.001 min
Response: 52461824
Conc: 481.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.013 min
Delta R.T.: 0.003 min
Response: 34376057
Conc: 483.85 ng/ml




#29 AR-1254-4

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 36730376
Conc: 493.77 ng/ml



#29 AR-1254-4


R.T.: 6.243 min
Delta R.T.: 0.003 min
Response: 21471276
Conc: 484.84 ng/ml



#30 AR-1254-5

R.T.: 7.660 min
Delta R.T.: 0.002 min
Response: 37552734
Conc: 494.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.663 min
Delta R.T.: 0.003 min
Response: 29319445
Conc: 483.45 ng/ml