

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
 Data File : PO089774.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2022 07:08
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 07:35:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 05 06:04:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.335	3.560	141.4E6	52036516	54.404	57.620
2) SA Decachlor...	10.093	8.590	63302872	39761088	58.888	50.038
Target Compounds						
26) L6 AR-1254-1	6.364	5.458	37511679	26815537	514.159	533.103
27) L6 AR-1254-2	6.583	5.606	55460085	22910693	521.572	524.280
28) L6 AR-1254-3	6.950	6.011	56091799	36278122	514.856	510.621
29) L6 AR-1254-4	7.238	6.241	39322286	22519031	528.614	508.502
30) L6 AR-1254-5	7.659	6.661	40317568	30379804	531.317	500.934



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100522\
Data File : PO089774.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2022 07:08
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 07:35:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 05 06:04:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.335 min
Delta R.T.: 0.000 min
Response: 141441101
Conc: 54.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.560 min
Delta R.T.: 0.004 min
Response: 52036516
Conc: 57.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.093 min
Delta R.T.: 0.001 min
Response: 63302872
Conc: 58.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.590 min
Delta R.T.: 0.002 min
Response: 39761088
Conc: 50.04 ng/ml



#26 AR-1254-1


R.T.: 6.364 min
Delta R.T.: 0.000 min
Response: 37511679
Conc: 514.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.458 min
Delta R.T.: 0.002 min
Response: 26815537
Conc: 533.10 ng/ml




#27 AR-1254-2

R.T.: 6.583 min
Delta R.T.: 0.000 min
Response: 55460085
Conc: 521.57 ng/ml



#27 AR-1254-2


R.T.: 5.606 min
Delta R.T.: 0.002 min
Response: 22910693
Conc: 524.28 ng/ml



#28 AR-1254-3


R.T.: 6.950 min
Delta R.T.: 0.000 min
Response: 56091799
Conc: 514.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.011 min
Delta R.T.: 0.002 min
Response: 36278122
Conc: 510.62 ng/ml




#29 AR-1254-4

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 39322286
Conc: 528.61 ng/ml



#29 AR-1254-4


R.T.: 6.241 min
Delta R.T.: 0.001 min
Response: 22519031
Conc: 508.50 ng/ml



#30 AR-1254-5

R.T.: 7.659 min
Delta R.T.: 0.000 min
Response: 40317568
Conc: 531.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.661 min
Delta R.T.: 0.001 min
Response: 30379804
Conc: 500.93 ng/ml