

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100621\
 Data File : PO081827.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2021 2:39
 Operator : AJ\MA
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 03:42:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 07 03:40:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.990	3.981	2104966	662061	25.280	24.873
2) SA Decachlor...	11.111	9.316	1561247	662075	25.759	26.311
Target Compounds						
21) L5 AR-1248-1	6.322	5.245	433235	157219	259.932	248.244
22) L5 AR-1248-2	6.619	5.510	572108	223992	262.746	263.056
23) L5 AR-1248-3	6.838	5.554	643560	237227	260.504	274.113
24) L5 AR-1248-4	7.270	5.738	675987	261218	259.255	262.101
25) L5 AR-1248-5	7.310	6.163	650047	250128	260.787	259.866



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100621\
Data File : PO081827.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Oct 2021 2:39
Operator : AJ\MA
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 03:42:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 07 03:40:34 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.990 min
Delta R.T.: -0.002 min
Response: 2104966
Conc: 25.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.981 min
Delta R.T.: -0.002 min
Response: 662061
Conc: 24.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.111 min
Delta R.T.: -0.002 min
Response: 1561247
Conc: 25.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.316 min
Delta R.T.: -0.002 min
Response: 662075
Conc: 26.31 ng/ml



#21 AR-1248-1


R.T.: 6.322 min
Delta R.T.: -0.003 min
Response: 433235
Conc: 259.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 5.245 min
Delta R.T.: -0.002 min
Response: 157219
Conc: 248.24 ng/ml




#22 AR-1248-2

R.T.: 6.619 min
Delta R.T.: -0.002 min
Response: 572108
Conc: 262.75 ng/ml



#22 AR-1248-2


R.T.: 5.510 min
Delta R.T.: -0.002 min
Response: 223992
Conc: 263.06 ng/ml



#23 AR-1248-3


R.T.: 6.838 min
Delta R.T.: -0.003 min
Response: 643560
Conc: 260.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.554 min
Delta R.T.: -0.002 min
Response: 237227
Conc: 274.11 ng/ml




#24 AR-1248-4

R.T.: 7.270 min
Delta R.T.: -0.002 min
Response: 675987
Conc: 259.26 ng/ml



#24 AR-1248-4


R.T.: 5.738 min
Delta R.T.: -0.003 min
Response: 261218
Conc: 262.10 ng/ml



#25 AR-1248-5

R.T.: 7.310 min
Delta R.T.: -0.002 min
Response: 650047
Conc: 260.79 ng/ml

Instrument :
ECD_O
ClientSampleId :



#25 AR-1248-5

R.T.: 6.163 min
Delta R.T.: -0.002 min
Response: 250128
Conc: 259.87 ng/ml