

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100622\
 Data File : PO089795.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2022 18:30
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 19:02:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 05 06:04:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.337	3.562	129.7E6	50586955	49.887	56.015
2) SA Decachlor...	10.097	8.593	58563301	40219978	54.479	50.616
Target Compounds						
26) L6 AR-1254-1	6.367	5.460	35945561	25648342	492.693	509.898
27) L6 AR-1254-2	6.586	5.609	52494847	22265073	493.686	509.506
28) L6 AR-1254-3	6.953	6.014	52574743	35657034	482.574	501.880
29) L6 AR-1254-4	7.240	6.244	37493202	22295134	504.025	503.447
30) L6 AR-1254-5	7.661	6.664	38476589	30720109	507.056	506.545



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100622\
Data File : PO089795.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2022 18:30
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 19:02:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 05 06:04:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.337 min
Delta R.T.: 0.002 min
Response: 129696769
Conc: 49.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.562 min
Delta R.T.: 0.006 min
Response: 50586955
Conc: 56.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.097 min
Delta R.T.: 0.005 min
Response: 58563301
Conc: 54.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.593 min
Delta R.T.: 0.005 min
Response: 40219978
Conc: 50.62 ng/ml



#26 AR-1254-1


R.T.: 6.367 min
Delta R.T.: 0.003 min
Response: 35945561
Conc: 492.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.460 min
Delta R.T.: 0.005 min
Response: 25648342
Conc: 509.90 ng/ml




#27 AR-1254-2

R.T.: 6.586 min
Delta R.T.: 0.003 min
Response: 52494847
Conc: 493.69 ng/ml



#27 AR-1254-2


R.T.: 5.609 min
Delta R.T.: 0.005 min
Response: 22265073
Conc: 509.51 ng/ml



#28 AR-1254-3


R.T.: 6.953 min
Delta R.T.: 0.003 min
Response: 52574743
Conc: 482.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.014 min
Delta R.T.: 0.005 min
Response: 35657034
Conc: 501.88 ng/ml




#29 AR-1254-4

R.T.: 7.240 min
Delta R.T.: 0.003 min
Response: 37493202
Conc: 504.03 ng/ml



#29 AR-1254-4


R.T.: 6.244 min
Delta R.T.: 0.004 min
Response: 22295134
Conc: 503.45 ng/ml



#30 AR-1254-5

R.T.: 7.661 min
Delta R.T.: 0.002 min
Response: 38476589
Conc: 507.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.664 min
Delta R.T.: 0.004 min
Response: 30720109
Conc: 506.54 ng/ml