

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100719\
 Data File : P0061760.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2019 20:44
 Operator : HP/AJ
 Sample : K5231-04
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 20190766-SS-COMPOSITE-6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 02:54:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.359	3.519	1023053	764244	24.944	23.632
2)	SA Decachlor...	10.019	8.382	567079	381026	12.177	11.769
Target Compounds							
3)	L1 AR-1016-1	5.547	0.000	5730	0	3.125	N.D. #
4)	L1 AR-1016-2	5.547	0.000	5730	0	2.217	N.D. #
5)	L1 AR-1016-3	5.604	4.798	6180	9019	3.847	9.004 #
6)	L1 AR-1016-4	5.724	4.837	533	7846	0.399	9.054 #
7)	L1 AR-1016-5	6.040	5.003	3665	8798	2.672	7.708 #
8)	L2 AR-1221-1	4.519	0.000	1545	0	3.342	N.D. #
9)	L2 AR-1221-2	4.663	3.807	13940	11584	38.828	37.564
10)	L2 AR-1221-3	4.732	0.000	4229	0	3.715	N.D. #
11)	L3 AR-1232-1	4.732	0.000	4229	0	3.024	N.D. #
12)	L3 AR-1232-2	5.256	0.000	5683	0	7.243	N.D. #
13)	L3 AR-1232-3	5.547	4.756	5730	3078	3.354	4.662 #
14)	L3 AR-1232-4	5.702	4.837	4906	7846	5.545	12.447 #
15)	L3 AR-1232-5	5.791	5.003	9481	8798	13.675	12.823
16)	L4 AR-1242-1	5.521	0.000	3510	0	2.471	N.D. #
17)	L4 AR-1242-2	5.547	0.000	5730	0	2.851	N.D. #
18)	L4 AR-1242-3	5.604	4.756	6180	3078	4.897	3.942
19)	L4 AR-1242-4	5.702	4.837	4906	7846	4.669	9.561 #
20)	L4 AR-1242-5	6.435	5.346	12937	5075	9.673	4.738 #
21)	L5 AR-1248-1	5.547	0.000	5730	0	4.878	N.D. #
22)	L5 AR-1248-2	5.835	4.837	10537	7846	6.246	6.659
23)	L5 AR-1248-3	6.040	4.837	3665	7846	1.961	6.321 #
24)	L5 AR-1248-4	6.435	5.003	12937	8798	5.475	5.853
25)	L5 AR-1248-5	6.435	5.346	12937	5075	5.706	3.308 #
26)	L6 AR-1254-1	6.370	5.346	11225	5075	4.920	2.314 #
27)	L6 AR-1254-2	6.590	5.488	21901	-37	6.062	N.D. #
28)	L6 AR-1254-3	6.957	5.884	17538	3340	4.488	1.049 #
29)	L6 AR-1254-4	7.237	0.000	22729	0	7.505	N.D. #
30)	L6 AR-1254-5	7.655	6.513	48187	4740	15.344	1.676 #
31)	L7 AR-1260-1	7.155	6.010	5345	1763	2.047	0.864 #
32)	L7 AR-1260-2	7.396	0.000	3583	0	1.124	N.D. #
33)	L7 AR-1260-3	7.758	6.414	5212	4819	2.170	2.139
34)	L7 AR-1260-4	8.000	0.000	42604	0	15.944	N.D. #
35)	L7 AR-1260-5	8.324	0.000	42968	0	8.684	N.D. #
36)	L8 AR-1262-1	7.731	0.000	48975	0	12.782	N.D. #
37)	L8 AR-1262-2	8.272	0.000	97192	0	15.656	N.D. #
38)	L8 AR-1262-3	8.570	7.325	87800	4971	21.125	2.639 #
39)	L8 AR-1262-4	8.674	7.413	28787	5866	9.119	1.746 #
40)	L8 AR-1262-5	9.298	0.000	18166	0	8.933	N.D. #
41)	L9 AR-1268-1	8.570	7.325	87800	4971	13.057	1.029 #
42)	L9 AR-1268-2	8.674	7.413	28787	5866	4.649	1.361 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100719\
 Data File : P0061760.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2019 20:44
 Operator : HP/AJ
 Sample : K5231-04
 Misc :
 ALS Vial : 31 Sample Multiplier: 1


Instrument :
 ECD_O
 ClientSampleId :
 20190766-SS-COMPOSITE-6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 02:54:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
43) L9	AR-1268-3	8.908	0.000	75696	0	14.807	N.D.	#
44) L9	AR-1268-4	9.318	0.000	6502	0	2.974	N.D.	#
45) L9	AR-1268-5	9.696	8.145	7761	984	0.517	0.098	#


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.



#1 Tetrachloro-m-xylene


R.T.: 4.359 min
Delta R.T.: -0.004 min
Response: 1023053
Conc: 24.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#1 Tetrachloro-m-xylene

R.T.: 3.519 min
Delta R.T.: 0.000 min
Response: 764244
Conc: 23.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.019 min
Delta R.T.: -0.017 min
Response: 567079
Conc: 12.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.382 min
Delta R.T.: -0.027 min
Response: 381026
Conc: 11.77 ng/ml



#3 AR-1016-1


R.T.: 5.547 min
Delta R.T.: -0.003 min
Response: 5730
Conc: 3.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.590 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.547 min
Delta R.T.: -0.023 min
Response: 5730
Conc: 2.22 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.610 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.604 min
Delta R.T.: -0.036 min
Response: 6180
Conc: 3.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#5 AR-1016-3

R.T.: 4.798 min
Delta R.T.: 0.018 min
Response: 9019
Conc: 9.00 ng/ml




#6 AR-1016-4

R.T.: 5.724 min
Delta R.T.: -0.006 min
Response: 533
Conc: 0.40 ng/ml



#6 AR-1016-4


R.T.: 4.837 min
Delta R.T.: 0.017 min
Response: 7846
Conc: 9.05 ng/ml



#7 AR-1016-5


R.T.: 6.040 min
Delta R.T.: 0.010 min
Response: 3665
Conc: 2.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#7 AR-1016-5

R.T.: 5.003 min
Delta R.T.: -0.027 min
Response: 8798
Conc: 7.71 ng/ml




#8 AR-1221-1

R.T.: 4.519 min
Delta R.T.: -0.055 min
Response: 1545
Conc: 3.34 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.735 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.663 min
Delta R.T.: 0.003 min
Response: 13940
Conc: 38.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#9 AR-1221-2

R.T.: 3.807 min
Delta R.T.: -0.009 min
Response: 11584
Conc: 37.56 ng/ml




#10 AR-1221-3

R.T.: 4.732 min
Delta R.T.: -0.004 min
Response: 4229
Conc: 3.71 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.891 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.732 min
Delta R.T.: -0.003 min
Response: 4229
Conc: 3.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.890 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.256 min
Delta R.T.: -0.005 min
Response: 5683
Conc: 7.24 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.595 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.547 min
Delta R.T.: -0.003 min
Response: 5730
Conc: 3.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#13 AR-1232-3

R.T.: 4.756 min
Delta R.T.: -0.009 min
Response: 3078
Conc: 4.66 ng/ml




#14 AR-1232-4

R.T.: 5.702 min
Delta R.T.: -0.008 min
Response: 4906
Conc: 5.54 ng/ml



#14 AR-1232-4


R.T.: 4.837 min
Delta R.T.: -0.010 min
Response: 7846
Conc: 12.45 ng/ml



#15 AR-1232-5


R.T.: 5.791 min
Delta R.T.: -0.008 min
Response: 9481
Conc: 13.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#15 AR-1232-5

R.T.: 5.003 min
Delta R.T.: -0.010 min
Response: 8798
Conc: 12.82 ng/ml




#16 AR-1242-1

R.T.: 5.521 min
Delta R.T.: -0.009 min
Response: 3510
Conc: 2.47 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.570 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.547 min
Delta R.T.: -0.003 min
Response: 5730
Conc: 2.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.590 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.604 min
Delta R.T.: -0.006 min
Response: 6180
Conc: 4.90 ng/ml



#18 AR-1242-3


R.T.: 4.756 min
Delta R.T.: -0.004 min
Response: 3078
Conc: 3.94 ng/ml



#19 AR-1242-4


R.T.: 5.702 min
Delta R.T.: -0.008 min
Response: 4906
Conc: 4.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#19 AR-1242-4

R.T.: 4.837 min
Delta R.T.: -0.003 min
Response: 7846
Conc: 9.56 ng/ml




#20 AR-1242-5

R.T.: 6.435 min
Delta R.T.: -0.005 min
Response: 12937
Conc: 9.67 ng/ml



#20 AR-1242-5


R.T.: 5.346 min
Delta R.T.: -0.004 min
Response: 5075
Conc: 4.74 ng/ml



#21 AR-1248-1


R.T.: 5.547 min
Delta R.T.: -0.003 min
Response: 5730
Conc: 4.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.590 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.835 min
Delta R.T.: 0.015 min
Response: 10537
Conc: 6.25 ng/ml



#22 AR-1248-2


R.T.: 4.837 min
Delta R.T.: 0.017 min
Response: 7846
Conc: 6.66 ng/ml



#23 AR-1248-3


R.T.: 6.040 min
Delta R.T.: 0.010 min
Response: 3665
Conc: 1.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#23 AR-1248-3

R.T.: 4.837 min
Delta R.T.: -0.023 min
Response: 7846
Conc: 6.32 ng/ml




#24 AR-1248-4

R.T.: 6.435 min
Delta R.T.: 0.005 min
Response: 12937
Conc: 5.48 ng/ml



#24 AR-1248-4


R.T.: 5.003 min
Delta R.T.: -0.027 min
Response: 8798
Conc: 5.85 ng/ml



#25 AR-1248-5


R.T.: 6.435 min
Delta R.T.: -0.035 min
Response: 12937
Conc: 5.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#25 AR-1248-5

R.T.: 5.346 min
Delta R.T.: -0.064 min
Response: 5075
Conc: 3.31 ng/ml




#26 AR-1254-1

R.T.: 6.370 min
Delta R.T.: -0.011 min
Response: 11225
Conc: 4.92 ng/ml



#26 AR-1254-1


R.T.: 5.346 min
Delta R.T.: -0.011 min
Response: 5075
Conc: 2.31 ng/ml



#27 AR-1254-2


R.T.: 6.590 min
Delta R.T.: -0.009 min
Response: 21901
Conc: 6.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#27 AR-1254-2

R.T.: 5.488 min
Delta R.T.: -0.013 min
Response: -37
Conc: N.D.




#28 AR-1254-3

R.T.: 6.957 min
Delta R.T.: -0.007 min
Response: 17538
Conc: 4.49 ng/ml



#28 AR-1254-3


R.T.: 5.884 min
Delta R.T.: -0.011 min
Response: 3340
Conc: 1.05 ng/ml



#29 AR-1254-4


R.T.: 7.237 min
Delta R.T.: -0.012 min
Response: 22729
Conc: 7.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.119 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.655 min
Delta R.T.: -0.012 min
Response: 48187
Conc: 15.34 ng/ml



#30 AR-1254-5


R.T.: 6.513 min
Delta R.T.: -0.017 min
Response: 4740
Conc: 1.68 ng/ml



#31 AR-1260-1


R.T.: 7.155 min
Delta R.T.: 0.005 min
Response: 5345
Conc: 2.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#31 AR-1260-1

R.T.: 6.010 min
Delta R.T.: -0.030 min
Response: 1763
Conc: 0.86 ng/ml




#32 AR-1260-2

R.T.: 7.396 min
Delta R.T.: -0.004 min
Response: 3583
Conc: 1.12 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 6.220 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.758 min
Delta R.T.: -0.002 min
Response: 5212
Conc: 2.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#33 AR-1260-3

R.T.: 6.414 min
Delta R.T.: 0.043 min
Response: 4819
Conc: 2.14 ng/ml




#34 AR-1260-4

R.T.: 8.000 min
Delta R.T.: 0.010 min
Response: 42604
Conc: 15.94 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.840 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.324 min
Delta R.T.: 0.024 min
Response: 42968
Conc: 8.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.070 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.731 min
Delta R.T.: -0.008 min
Response: 48975
Conc: 12.78 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.619 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.272 min
Delta R.T.: -0.007 min
Response: 97192
Conc: 15.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.061 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.570 min
Delta R.T.: -0.011 min
Response: 87800
Conc: 21.12 ng/ml



#38 AR-1262-3


R.T.: 7.325 min
Delta R.T.: -0.015 min
Response: 4971
Conc: 2.64 ng/ml



#39 AR-1262-4


R.T.: 8.674 min
Delta R.T.: 0.004 min
Response: 28787
Conc: 9.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#39 AR-1262-4

R.T.: 7.413 min
Delta R.T.: 0.009 min
Response: 5866
Conc: 1.75 ng/ml




#40 AR-1262-5

R.T.: 9.298 min
Delta R.T.: -0.009 min
Response: 18166
Conc: 8.93 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 7.891 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.570 min
Delta R.T.: -0.012 min
Response: 87800
Conc: 13.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#41 AR-1268-1

R.T.: 7.325 min
Delta R.T.: -0.016 min
Response: 4971
Conc: 1.03 ng/ml




#42 AR-1268-2

R.T.: 8.674 min
Delta R.T.: 0.000 min
Response: 28787
Conc: 4.65 ng/ml



#42 AR-1268-2


R.T.: 7.413 min
Delta R.T.: 0.008 min
Response: 5866
Conc: 1.36 ng/ml



#43 AR-1268-3


R.T.: 8.908 min
Delta R.T.: 0.015 min
Response: 75696
Conc: 14.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.607 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.318 min
Delta R.T.: 0.009 min
Response: 6502
Conc: 2.97 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 7.891 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.696 min
Delta R.T.: -0.012 min
Response: 7761
Conc: 0.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
20190766-SS-COMPOSITE-6



#45 AR-1268-5

R.T.: 8.145 min
Delta R.T.: -0.018 min
Response: 984
Conc: 0.10 ng/ml