

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100719\
 Data File : P0061753.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2019 18:50
 Operator : HP/AJ
 Sample : K5213-05
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S-3(0-5)

Manual Integrations
 APPROVED

Ankita
 10/8/2019 11:06:45 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 02:49:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.361	3.522	775029	433022	18.897	13.390m#
2)	SA Decachlor...	10.022	8.383	640545	400322	13.754	12.365
Target Compounds							
26)	L6 AR-1254-1	6.373	5.349	2170603	1416099	951.437	645.650m#
27)	L6 AR-1254-2	6.591	5.493	3722026	1543434	1030.156	787.269m
28)	L6 AR-1254-3	6.956	5.886	4499340	2784117	1151.507	874.476m
29)	L6 AR-1254-4	7.241	6.110	2618687	1349413	864.645	665.113m
30)	L6 AR-1254-5	7.658	6.521	3888746	2863947	1238.293	1012.545

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100719\
Data File : PO061753.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Oct 2019 18:50
Operator : HP/AJ
Sample : K5213-05
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
S-3(0-5)

Manual Integrations
APPROVED

Ankita
10/8/2019 11:06:45 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 02:49:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

