

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100719\
 Data File : P0061730.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2019 8:37
 Operator : HP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 09:27:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 01 15:34:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.361	3.520	2222003	1660002	54.176	51.331
2) SA Decachlor...	10.020	8.382	2334027	1548701	50.118	47.835
Target Compounds						
3) L1 AR-1016-1	5.524	4.570	901244	634078	491.477	471.552
4) L1 AR-1016-2	5.546	4.588	1265782	875420	489.699	472.342
5) L1 AR-1016-3	5.608	4.758	781640	485615	486.566	484.781
6) L1 AR-1016-4	5.706	4.800	650711	410358	486.715	473.546
7) L1 AR-1016-5	5.998	5.006	665662	528106	485.380	462.667
31) L7 AR-1260-1	7.119	6.013	1281209	922536	490.783	452.268
32) L7 AR-1260-2	7.376	6.198	1558321	1093838	488.809	455.436
33) L7 AR-1260-3	7.733	6.348	1251869	1024459	521.207	454.686
34) L7 AR-1260-4	7.958	6.811	1429220	864779	534.878	468.408
35) L7 AR-1260-5	8.272	7.051	2605455	1871522	526.554	480.663



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100719\
Data File : PO061730.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Oct 2019 8:37
Operator : HP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 09:27:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 01 15:34:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.361 min
Delta R.T.: -0.004 min
Response: 2222003
Conc: 54.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.520 min
Delta R.T.: -0.006 min
Response: 1660002
Conc: 51.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.020 min
Delta R.T.: -0.004 min
Response: 2334027
Conc: 50.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.382 min
Delta R.T.: -0.012 min
Response: 1548701
Conc: 47.84 ng/ml



#3 AR-1016-1


R.T.: 5.524 min
Delta R.T.: -0.005 min
Response: 901244
Conc: 491.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.570 min
Delta R.T.: -0.007 min
Response: 634078
Conc: 471.55 ng/ml




#4 AR-1016-2

R.T.: 5.546 min
Delta R.T.: -0.005 min
Response: 1265782
Conc: 489.70 ng/ml



#4 AR-1016-2


R.T.: 4.588 min
Delta R.T.: -0.007 min
Response: 875420
Conc: 472.34 ng/ml



#5 AR-1016-3


R.T.: 5.608 min
Delta R.T.: -0.005 min
Response: 781640
Conc: 486.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.758 min
Delta R.T.: -0.007 min
Response: 485615
Conc: 484.78 ng/ml




#6 AR-1016-4

R.T.: 5.706 min
Delta R.T.: -0.005 min
Response: 650711
Conc: 486.72 ng/ml



#6 AR-1016-4


R.T.: 4.800 min
Delta R.T.: -0.008 min
Response: 410358
Conc: 473.55 ng/ml



#7 AR-1016-5


R.T.: 5.998 min
Delta R.T.: -0.005 min
Response: 665662
Conc: 485.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.006 min
Delta R.T.: -0.007 min
Response: 528106
Conc: 462.67 ng/ml




#31 AR-1260-1

R.T.: 7.119 min
Delta R.T.: -0.006 min
Response: 1281209
Conc: 490.78 ng/ml



#31 AR-1260-1


R.T.: 6.013 min
Delta R.T.: -0.009 min
Response: 922536
Conc: 452.27 ng/ml



#32 AR-1260-2


R.T.: 7.376 min
Delta R.T.: -0.005 min
Response: 1558321
Conc: 488.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.198 min
Delta R.T.: -0.009 min
Response: 1093838
Conc: 455.44 ng/ml




#33 AR-1260-3

R.T.: 7.733 min
Delta R.T.: -0.005 min
Response: 1251869
Conc: 521.21 ng/ml



#33 AR-1260-3


R.T.: 6.348 min
Delta R.T.: -0.009 min
Response: 1024459
Conc: 454.69 ng/ml



#34 AR-1260-4


R.T.: 7.958 min
Delta R.T.: -0.005 min
Response: 1429220
Conc: 534.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.811 min
Delta R.T.: -0.009 min
Response: 864779
Conc: 468.41 ng/ml



#35 AR-1260-5

R.T.: 8.272 min
Delta R.T.: -0.005 min
Response: 2605455
Conc: 526.55 ng/ml



#35 AR-1260-5

R.T.: 7.051 min
Delta R.T.: -0.009 min
Response: 1871522
Conc: 480.66 ng/ml