

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100721\
 Data File : P0081850.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2021 10:38
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 23:09:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 07 09:43:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.992	3.981	1769027	546516	21.318	21.598
2)	SA Decachlor...	11.114	9.315	1371879	595759	23.129	25.219
Target Compounds							
8)	L2 AR-1221-1	5.235	0.000	6116	0	6.405	N.D. #
9)	L2 AR-1221-2	5.347	0.000	24526	0	37.989	N.D. #
24)	L5 AR-1248-4	7.250f	0.000	29328	0	11.207	N.D. #
26)	L6 AR-1254-1	7.250	0.000	29328	0	10.720	N.D. #
28)	L6 AR-1254-3	7.875f	0.000	24106	0	5.908	N.D. #
30)	L6 AR-1254-5	8.593	0.000	8460	0	2.591	N.D. #
32)	L7 AR-1260-2	8.286	0.000	2057	0	0.557	N.D. #
33)	L7 AR-1260-3	8.676f	0.000	3602	0	1.457	N.D. #
36)	L8 AR-1262-1	8.676f	0.000	3602	0	0.916	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100721\
Data File : PO081850.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Oct 2021 10:38
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 23:09:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 07 09:43:36 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.992 min
Delta R.T.: 0.002 min
Response: 1769027
Conc: 21.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.981 min
Delta R.T.: 0.000 min
Response: 546516
Conc: 21.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.114 min
Delta R.T.: 0.004 min
Response: 1371879
Conc: 23.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.315 min
Delta R.T.: -0.005 min
Response: 595759
Conc: 25.22 ng/ml



#8 AR-1221-1


R.T.: 5.235 min
Delta R.T.: -0.009 min
Response: 6116
Conc: 6.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.243 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.347 min
Delta R.T.: 0.004 min
Response: 24526
Conc: 37.99 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.342 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 7.250 min
Delta R.T.: -0.021 min
Response: 29328
Conc: 11.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.739 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 7.250 min
Delta R.T.: 0.010 min
Response: 29328
Conc: 10.72 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 6.121 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.875 min
Delta R.T.: 0.020 min
Response: 24106
Conc: 5.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.702 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 8.593 min
Delta R.T.: 0.009 min
Response: 8460
Conc: 2.59 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 7.375 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.286 min
Delta R.T.: -0.003 min
Response: 2057
Conc: 0.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.042 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.676 min
Delta R.T.: 0.018 min
Response: 3602
Conc: 1.46 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.197 min
Response: 0
Conc: N.D.



#36 AR-1262-1

R.T.: 8.676 min
Delta R.T.: 0.018 min
Response: 3602
Conc: 0.92 ng/ml

Instrument :
ECD_O
ClientSampleId :



#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.375 min
Response: 0
Conc: N.D.