

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100721\
 Data File : PO081847.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2021 9:46
 Operator : AJ\MA
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 23:07:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 07 09:43:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.991	3.980	4343037	1301848	52.337	51.448
2)	SA Decachlor...	11.113	9.314	3124282	1279164	52.673	54.149
Target Compounds							
16)	L4 AR-1242-1	6.324	5.244	1079760	398584	525.048	516.206
17)	L4 AR-1242-2	6.347	5.264	1540806	542937	525.645	514.244
18)	L4 AR-1242-3	6.414	5.455	979892	292882	537.341	518.496
19)	L4 AR-1242-4	6.522	5.553	786105	310319	551.672	497.878
20)	L4 AR-1242-5	7.311	6.119	830071	384426	568.628	526.834



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100721\
Data File : PO081847.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Oct 2021 9:46
Operator : AJ\MA
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 23:07:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 07 09:43:36 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.991 min
Delta R.T.: 0.000 min
Response: 4343037
Conc: 52.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.980 min
Delta R.T.: -0.001 min
Response: 1301848
Conc: 51.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.113 min
Delta R.T.: 0.004 min
Response: 3124282
Conc: 52.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.314 min
Delta R.T.: -0.006 min
Response: 1279164
Conc: 54.15 ng/ml



#16 AR-1242-1


R.T.: 6.324 min
Delta R.T.: 0.000 min
Response: 1079760
Conc: 525.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.244 min
Delta R.T.: -0.003 min
Response: 398584
Conc: 516.21 ng/ml




#17 AR-1242-2

R.T.: 6.347 min
Delta R.T.: 0.000 min
Response: 1540806
Conc: 525.65 ng/ml



#17 AR-1242-2


R.T.: 5.264 min
Delta R.T.: -0.003 min
Response: 542937
Conc: 514.24 ng/ml



#18 AR-1242-3


R.T.: 6.414 min
Delta R.T.: 0.000 min
Response: 979892
Conc: 537.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.455 min
Delta R.T.: -0.004 min
Response: 292882
Conc: 518.50 ng/ml




#19 AR-1242-4

R.T.: 6.522 min
Delta R.T.: 0.000 min
Response: 786105
Conc: 551.67 ng/ml



#19 AR-1242-4


R.T.: 5.553 min
Delta R.T.: -0.004 min
Response: 310319
Conc: 497.88 ng/ml



#20 AR-1242-5

R.T.: 7.311 min
Delta R.T.: 0.000 min
Response: 830071
Conc: 568.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 6.119 min
Delta R.T.: -0.004 min
Response: 384426
Conc: 526.83 ng/ml