

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100824\
 Data File : PO107001.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Oct 2024 19:22
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 02:17:27 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 02:15:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.375	3.647	58749723	18364169	5.085	4.595
2) SA Decachlor...	10.056	8.645	12009757	13456130	4.927	4.941
Target Compounds						
26) L6 AR-1254-1	6.374	5.532	16185565	8728836	53.289	51.606
27) L6 AR-1254-2	6.593	5.678	22415575	7888149	52.941	52.570
28) L6 AR-1254-3	6.957	6.082	20266831	11797337	51.117	50.460
29) L6 AR-1254-4	7.242	6.309	14257530	6317419	54.187	49.112
30) L6 AR-1254-5	7.661	6.727	15867452	9925484	58.518	48.834



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100824\
Data File : PO107001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Oct 2024 19:22
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 02:17:27 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 02:15:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.375 min
Delta R.T.: 0.000 min
Response: 58749723
Conc: 5.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: 0.000 min
Response: 18364169
Conc: 4.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.056 min
Delta R.T.: 0.000 min
Response: 12009757
Conc: 4.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.645 min
Delta R.T.: 0.000 min
Response: 13456130
Conc: 4.94 ng/ml



#26 AR-1254-1


R.T.: 6.374 min
Delta R.T.: 0.000 min
Response: 16185565
Conc: 53.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 5.532 min
Delta R.T.: 0.000 min
Response: 8728836
Conc: 51.61 ng/ml




#27 AR-1254-2

R.T.: 6.593 min
Delta R.T.: 0.000 min
Response: 22415575
Conc: 52.94 ng/ml



#27 AR-1254-2


R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 7888149
Conc: 52.57 ng/ml



#28 AR-1254-3


R.T.: 6.957 min
Delta R.T.: 0.000 min
Response: 20266831
Conc: 51.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.082 min
Delta R.T.: 0.000 min
Response: 11797337
Conc: 50.46 ng/ml




#29 AR-1254-4

R.T.: 7.242 min
Delta R.T.: 0.000 min
Response: 14257530
Conc: 54.19 ng/ml



#29 AR-1254-4


R.T.: 6.309 min
Delta R.T.: 0.000 min
Response: 6317419
Conc: 49.11 ng/ml



#30 AR-1254-5

R.T.: 7.661 min
Delta R.T.: 0.000 min
Response: 15867452
Conc: 58.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 6.727 min
Delta R.T.: 0.000 min
Response: 9925484
Conc: 48.83 ng/ml