

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0100923\
 Data File : P0098604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 03:13
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 05:24:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.480	3.635	40006391	15490641	20.835	19.889
2)	SA Decachlor...	10.296	8.643	26738054	11746076	22.027	21.924
Target Compounds							
3)	L1 AR-1016-1	5.657	4.740	35947	50159	0.665	2.102 #
4)	L1 AR-1016-2	5.678	4.740	55120	50159	0.679	1.473 #
5)	L1 AR-1016-3	5.752	4.919	22883	25059	0.457	1.393 #
6)	L1 AR-1016-4	5.843	4.974	7219	69215	0.185	4.617 #
7)	L1 AR-1016-5	6.146	5.182	-22185	158254	N.D.	8.149 #
8)	L2 AR-1221-1	4.690	3.866	65378	61783	2.904	6.725 #
9)	L2 AR-1221-2	4.791	3.936	889259	241081	52.703	36.256 #
10)	L2 AR-1221-3	4.846	4.013	56494	9766	1.155	0.484 #
11)	L3 AR-1232-1	4.846	4.013	56494	9766	1.383	0.581 #
12)	L3 AR-1232-2	5.393	4.740	41101	50159	1.900	3.338 #
13)	L3 AR-1232-3	5.678	4.919	55120	25059	1.551	3.202 #
14)	L3 AR-1232-4	5.843	4.995	7219	101435	0.425	13.490 #
15)	L3 AR-1232-5	5.946	5.182	84546	158254	5.938	19.562 #
16)	L4 AR-1242-1	5.657	4.740	35947	50159	0.789	2.396 #
17)	L4 AR-1242-2	5.678	4.740	55120	50159	0.805	1.725 #
18)	L4 AR-1242-3	5.752	4.919	22883	25059	0.540	1.603 #
19)	L4 AR-1242-4	5.843	4.995	7219	101435	0.220	6.124 #
20)	L4 AR-1242-5	6.597	5.535	340255	232686	9.606	12.538 #
21)	L5 AR-1248-1	5.657	4.740	35947	50159	1.062	3.201 #
22)	L5 AR-1248-2	5.946	4.974	84546	69215	1.633	3.145 #
23)	L5 AR-1248-3	6.146	4.995	-22185	101435	N.D.	4.320 #
24)	L5 AR-1248-4	6.553	5.182	169678	158254	2.993	5.933 #
25)	L5 AR-1248-5	6.597	5.562	340255	110014	5.753	4.662
26)	L6 AR-1254-1	6.521	5.535	298476	232686	4.725	6.068 #
27)	L6 AR-1254-2	6.745	5.671	73740	197381	0.768	5.733 #
28)	L6 AR-1254-3	7.107	6.091	391737	101908	4.163	1.961 #
29)	L6 AR-1254-4	7.409	6.307	230141	33318	3.953	1.214 #
30)	L6 AR-1254-5	7.805	6.727	62127	48015	0.889	1.108
31)	L7 AR-1260-1	7.259	6.204	48613	47415	0.669	1.292 #
32)	L7 AR-1260-2	7.531	6.402	251690	13006	3.057	0.310 #
33)	L7 AR-1260-3	7.896	6.547	76320	30343	1.234	0.758 #
34)	L7 AR-1260-4	8.104	7.031	652050	13721	9.673	0.426 #
35)	L7 AR-1260-5	8.435	7.247	240775	83872	1.853	1.197 #
36)	L8 AR-1262-1	7.896	6.761	76320	21691	0.878	0.479 #
37)	L8 AR-1262-2	8.445	7.031	203022	13721	1.439	0.340 #
38)	L8 AR-1262-3	8.774	7.581	83890	289558	0.850	9.327 #
39)	L8 AR-1262-4	8.852	7.605	55698	116283	0.726	2.146 #
40)	L8 AR-1262-5	9.510	8.116	27366	68181	0.564	2.761 #
41)	L9 AR-1268-1	8.774	7.581	83890	289558	0.483	3.169 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100923\
 Data File : PO098604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 03:13
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 05:24:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.852	7.605	55698	116283	0.357	1.500 #
43) L9	AR-1268-3	9.091	7.816	17570	26372	0.129	0.398 #
44) L9	AR-1268-4	9.546	8.116	127320	68181	2.390	2.519
45) L9	AR-1268-5	9.968	8.400	42812	40688	0.097	0.206 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100923\
Data File : PO098604.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 03:13
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 05:24:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.480 min
Delta R.T.: -0.005 min
Response: 40006391
Conc: 20.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.635 min
Delta R.T.: -0.002 min
Response: 15490641
Conc: 19.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.296 min
Delta R.T.: 0.002 min
Response: 26738054
Conc: 22.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.643 min
Delta R.T.: -0.009 min
Response: 11746076
Conc: 21.92 ng/ml



#3 AR-1016-1


R.T.: 5.657 min
Delta R.T.: -0.001 min
Response: 35947
Conc: 0.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.740 min
Delta R.T.: 0.017 min
Response: 50159
Conc: 2.10 ng/ml




#4 AR-1016-2

R.T.: 5.678 min
Delta R.T.: -0.003 min
Response: 55120
Conc: 0.68 ng/ml



#4 AR-1016-2


R.T.: 4.740 min
Delta R.T.: -0.002 min
Response: 50159
Conc: 1.47 ng/ml



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: 0.008 min
Response: 22883
Conc: 0.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.919 min
Delta R.T.: 0.001 min
Response: 25059
Conc: 1.39 ng/ml




#6 AR-1016-4

R.T.: 5.843 min
Delta R.T.: 0.002 min
Response: 7219
Conc: 0.18 ng/ml



#6 AR-1016-4


R.T.: 4.974 min
Delta R.T.: 0.014 min
Response: 69215
Conc: 4.62 ng/ml



#7 AR-1016-5


R.T.: 6.146 min
Delta R.T.: 0.008 min
Response: -22185
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.182 min
Delta R.T.: 0.009 min
Response: 158254
Conc: 8.15 ng/ml




#8 AR-1221-1

R.T.: 4.690 min
Delta R.T.: -0.004 min
Response: 65378
Conc: 2.90 ng/ml



#8 AR-1221-1


R.T.: 3.866 min
Delta R.T.: 0.013 min
Response: 61783
Conc: 6.72 ng/ml



#9 AR-1221-2


R.T.: 4.791 min
Delta R.T.: 0.010 min
Response: 889259
Conc: 52.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.936 min
Delta R.T.: -0.003 min
Response: 241081
Conc: 36.26 ng/ml




#10 AR-1221-3

R.T.: 4.846 min
Delta R.T.: -0.012 min
Response: 56494
Conc: 1.16 ng/ml



#10 AR-1221-3


R.T.: 4.013 min
Delta R.T.: -0.001 min
Response: 9766
Conc: 0.48 ng/ml



#11 AR-1232-1


R.T.: 4.846 min
Delta R.T.: -0.012 min
Response: 56494
Conc: 1.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.013 min
Delta R.T.: -0.001 min
Response: 9766
Conc: 0.58 ng/ml




#12 AR-1232-2

R.T.: 5.393 min
Delta R.T.: 0.003 min
Response: 41101
Conc: 1.90 ng/ml



#12 AR-1232-2


R.T.: 4.740 min
Delta R.T.: -0.002 min
Response: 50159
Conc: 3.34 ng/ml



#13 AR-1232-3


R.T.: 5.678 min
Delta R.T.: -0.004 min
Response: 55120
Conc: 1.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.919 min
Delta R.T.: 0.001 min
Response: 25059
Conc: 3.20 ng/ml




#14 AR-1232-4

R.T.: 5.843 min
Delta R.T.: 0.000 min
Response: 7219
Conc: 0.43 ng/ml



#14 AR-1232-4


R.T.: 4.995 min
Delta R.T.: -0.007 min
Response: 101435
Conc: 13.49 ng/ml



#15 AR-1232-5


R.T.: 5.946 min
Delta R.T.: 0.012 min
Response: 84546
Conc: 5.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.182 min
Delta R.T.: 0.008 min
Response: 158254
Conc: 19.56 ng/ml




#16 AR-1242-1

R.T.: 5.657 min
Delta R.T.: -0.004 min
Response: 35947
Conc: 0.79 ng/ml



#16 AR-1242-1


R.T.: 4.740 min
Delta R.T.: 0.017 min
Response: 50159
Conc: 2.40 ng/ml



#17 AR-1242-2


R.T.: 5.678 min
Delta R.T.: -0.004 min
Response: 55120
Conc: 0.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.740 min
Delta R.T.: -0.002 min
Response: 50159
Conc: 1.73 ng/ml




#18 AR-1242-3

R.T.: 5.752 min
Delta R.T.: 0.007 min
Response: 22883
Conc: 0.54 ng/ml



#18 AR-1242-3


R.T.: 4.919 min
Delta R.T.: 0.000 min
Response: 25059
Conc: 1.60 ng/ml



#19 AR-1242-4


R.T.: 5.843 min
Delta R.T.: 0.000 min
Response: 7219
Conc: 0.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.995 min
Delta R.T.: -0.007 min
Response: 101435
Conc: 6.12 ng/ml




#20 AR-1242-5

R.T.: 6.597 min
Delta R.T.: 0.013 min
Response: 340255
Conc: 9.61 ng/ml



#20 AR-1242-5


R.T.: 5.535 min
Delta R.T.: 0.009 min
Response: 232686
Conc: 12.54 ng/ml



#21 AR-1248-1


R.T.: 5.657 min
Delta R.T.: -0.006 min
Response: 35947
Conc: 1.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.740 min
Delta R.T.: 0.018 min
Response: 50159
Conc: 3.20 ng/ml




#22 AR-1248-2

R.T.: 5.946 min
Delta R.T.: 0.008 min
Response: 84546
Conc: 1.63 ng/ml



#22 AR-1248-2


R.T.: 4.974 min
Delta R.T.: 0.015 min
Response: 69215
Conc: 3.15 ng/ml



#23 AR-1248-3


R.T.: 6.146 min
Delta R.T.: 0.004 min
Response: -22185
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.995 min
Delta R.T.: -0.005 min
Response: 101435
Conc: 4.32 ng/ml




#24 AR-1248-4

R.T.: 6.553 min
Delta R.T.: 0.004 min
Response: 169678
Conc: 2.99 ng/ml



#24 AR-1248-4


R.T.: 5.182 min
Delta R.T.: 0.009 min
Response: 158254
Conc: 5.93 ng/ml



#25 AR-1248-5


R.T.: 6.597 min
Delta R.T.: 0.010 min
Response: 340255
Conc: 5.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.562 min
Delta R.T.: -0.005 min
Response: 110014
Conc: 4.66 ng/ml




#26 AR-1254-1

R.T.: 6.521 min
Delta R.T.: -0.001 min
Response: 298476
Conc: 4.73 ng/ml



#26 AR-1254-1


R.T.: 5.535 min
Delta R.T.: 0.010 min
Response: 232686
Conc: 6.07 ng/ml



#27 AR-1254-2


R.T.: 6.745 min
Delta R.T.: 0.004 min
Response: 73740
Conc: 0.77 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.671 min
Delta R.T.: -0.002 min
Response: 197381
Conc: 5.73 ng/ml




#28 AR-1254-3

R.T.: 7.107 min
Delta R.T.: -0.002 min
Response: 391737
Conc: 4.16 ng/ml



#28 AR-1254-3


R.T.: 6.091 min
Delta R.T.: 0.013 min
Response: 101908
Conc: 1.96 ng/ml



#29 AR-1254-4


R.T.: 7.409 min
Delta R.T.: 0.013 min
Response: 230141
Conc: 3.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.307 min
Delta R.T.: 0.000 min
Response: 33318
Conc: 1.21 ng/ml




#30 AR-1254-5

R.T.: 7.805 min
Delta R.T.: -0.010 min
Response: 62127
Conc: 0.89 ng/ml



#30 AR-1254-5


R.T.: 6.727 min
Delta R.T.: 0.003 min
Response: 48015
Conc: 1.11 ng/ml



#31 AR-1260-1


R.T.: 7.259 min
Delta R.T.: -0.008 min
Response: 48613
Conc: 0.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.204 min
Delta R.T.: -0.005 min
Response: 47415
Conc: 1.29 ng/ml




#32 AR-1260-2

R.T.: 7.531 min
Delta R.T.: 0.007 min
Response: 251690
Conc: 3.06 ng/ml



#32 AR-1260-2


R.T.: 6.402 min
Delta R.T.: 0.005 min
Response: 13006
Conc: 0.31 ng/ml



#33 AR-1260-3


R.T.: 7.896 min
Delta R.T.: 0.011 min
Response: 76320
Conc: 1.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.547 min
Delta R.T.: -0.004 min
Response: 30343
Conc: 0.76 ng/ml




#34 AR-1260-4

R.T.: 8.104 min
Delta R.T.: -0.009 min
Response: 652050
Conc: 9.67 ng/ml



#34 AR-1260-4


R.T.: 7.031 min
Delta R.T.: 0.007 min
Response: 13721
Conc: 0.43 ng/ml



#35 AR-1260-5


R.T.: 8.435 min
Delta R.T.: -0.003 min
Response: 240775
Conc: 1.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.247 min
Delta R.T.: -0.019 min
Response: 83872
Conc: 1.20 ng/ml




#36 AR-1262-1

R.T.: 7.896 min
Delta R.T.: 0.006 min
Response: 76320
Conc: 0.88 ng/ml



#36 AR-1262-1


R.T.: 6.761 min
Delta R.T.: -0.002 min
Response: 21691
Conc: 0.48 ng/ml



#37 AR-1262-2


R.T.: 8.445 min
Delta R.T.: 0.002 min
Response: 203022
Conc: 1.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.031 min
Delta R.T.: 0.009 min
Response: 13721
Conc: 0.34 ng/ml




#38 AR-1262-3

R.T.: 8.774 min
Delta R.T.: 0.012 min
Response: 83890
Conc: 0.85 ng/ml



#38 AR-1262-3


R.T.: 7.581 min
Delta R.T.: 0.033 min
Response: 289558
Conc: 9.33 ng/ml



#39 AR-1262-4


R.T.: 8.852 min
Delta R.T.: 0.000 min
Response: 55698
Conc: 0.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.605 min
Delta R.T.: -0.006 min
Response: 116283
Conc: 2.15 ng/ml




#40 AR-1262-5

R.T.: 9.510 min
Delta R.T.: -0.018 min
Response: 27366
Conc: 0.56 ng/ml



#40 AR-1262-5


R.T.: 8.116 min
Delta R.T.: 0.008 min
Response: 68181
Conc: 2.76 ng/ml



#41 AR-1268-1


R.T.: 8.774 min
Delta R.T.: 0.012 min
Response: 83890
Conc: 0.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.581 min
Delta R.T.: 0.031 min
Response: 289558
Conc: 3.17 ng/ml




#42 AR-1268-2

R.T.: 8.852 min
Delta R.T.: -0.006 min
Response: 55698
Conc: 0.36 ng/ml



#42 AR-1268-2


R.T.: 7.605 min
Delta R.T.: -0.009 min
Response: 116283
Conc: 1.50 ng/ml



#43 AR-1268-3


R.T.: 9.091 min
Delta R.T.: 0.000 min
Response: 17570
Conc: 0.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.816 min
Delta R.T.: -0.005 min
Response: 26372
Conc: 0.40 ng/ml




#44 AR-1268-4

R.T.: 9.546 min
Delta R.T.: 0.018 min
Response: 127320
Conc: 2.39 ng/ml



#44 AR-1268-4


R.T.: 8.116 min
Delta R.T.: 0.007 min
Response: 68181
Conc: 2.52 ng/ml



#45 AR-1268-5

R.T.: 9.968 min
Delta R.T.: 0.012 min
Response: 42812
Conc: 0.10 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.400 min
Delta R.T.: 0.002 min
Response: 40688
Conc: 0.21 ng/ml