

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0100923\
 Data File : P0098612.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 05:28
 Operator : YP/AJ
 Sample : 04699-07
 Misc : AR1268 LOD 40 PPB
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 06:00:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.480	3.635	40932758	16145178	21.317	20.729
2)	SA Decachlor...	10.294	8.642	31732414	13667488	26.142	25.510
Target Compounds							
3)	L1 AR-1016-1	5.648	4.735	166000	89592	3.072	3.754
4)	L1 AR-1016-2	5.702	4.735	183753	89592	2.263	2.631
5)	L1 AR-1016-3	5.732	4.912	10528	35732	0.210	1.986 #
6)	L1 AR-1016-4	5.847	4.968	12086	47915	0.309	3.196 #
7)	L1 AR-1016-5	6.131	5.195	17863	86847	0.434	4.472 #
8)	L2 AR-1221-1	4.701	3.861	43463	58260	1.930	6.341 #
9)	L2 AR-1221-2	4.789	3.934	793031	258986	47.000	38.948
10)	L2 AR-1221-3	4.863	4.009	45966	45584	0.940	2.260 #
11)	L3 AR-1232-1	4.863	4.009	45966	45584	1.125	2.712 #
12)	L3 AR-1232-2	5.386	4.735	21040	89592	0.973	5.962 #
13)	L3 AR-1232-3	5.702	4.912	183753	35732	5.169	4.565
14)	L3 AR-1232-4	5.847	5.002	12086	26865	0.712	3.573 #
15)	L3 AR-1232-5	5.930	5.195	30338	86847	2.131	10.735 #
16)	L4 AR-1242-1	5.648	4.735	166000	89592	3.645	4.279
17)	L4 AR-1242-2	5.702	4.735	183753	89592	2.683	3.081
18)	L4 AR-1242-3	5.732	4.912	10528	35732	0.249	2.286 #
19)	L4 AR-1242-4	5.847	5.002	12086	26865	0.368	1.622 #
20)	L4 AR-1242-5	6.567	0.000	92920	0	2.623	N.D. #
21)	L5 AR-1248-1	5.648	4.735	166000	89592	4.905	5.717
22)	L5 AR-1248-2	5.940	4.968	25452	47915	0.491	2.177 #
23)	L5 AR-1248-3	6.150	5.002	39194	26865	0.694	1.144 #
24)	L5 AR-1248-4	6.551	5.195	110621	86847	1.951	3.256 #
25)	L5 AR-1248-5	6.606	0.000	383006	0	6.475	N.D. #
26)	L6 AR-1254-1	6.516	0.000	61281	0	0.970	N.D. #
27)	L6 AR-1254-2	6.740	5.638f	109533	508271	1.141	14.763 #
28)	L6 AR-1254-3	7.107	0.000	160987	0	1.711	N.D. #
29)	L6 AR-1254-4	7.404	6.278	82621	171271	1.419	6.240 #
30)	L6 AR-1254-5	7.800	6.696	263762	268300	3.776	6.194 #
31)	L7 AR-1260-1	7.266	6.218	14927	427717	0.205	11.657 #
32)	L7 AR-1260-2	7.539	6.379	617731	131841	7.502	3.142 #
33)	L7 AR-1260-3	7.885	6.561	1421444	391877	22.977	9.793 #
34)	L7 AR-1260-4	8.103	7.014	444182	979276	6.589	30.409 #
35)	L7 AR-1260-5	8.437	7.259	2306382	507931	17.752	7.250 #
36)	L8 AR-1262-1	7.885	6.758	1421444	796593	16.344	17.599
37)	L8 AR-1262-2	8.437	7.014	2306382	979276	16.351	24.285 #
38)	L8 AR-1262-3	8.753	7.542	7025661	3804036	71.159	122.538 #
39)	L8 AR-1262-4	8.850	7.607	6307862	3314045	82.196	61.175 #
40)	L8 AR-1262-5	9.520	8.102	2173468	1043109	44.806	42.239
41)	L9 AR-1268-1	8.753	7.542	7025661	3804036	40.460	41.629

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100923\
Data File : PO098612.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 05:28
Operator : YP/AJ
Sample : 04699-07
Misc : AR1268 LOD 40 PPB
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 06:00:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.850	7.607	6307862	3314045	40.454	42.741
43)	L9 AR-1268-3	9.084	7.813	5526444	2773414	40.419	41.854
44)	L9 AR-1268-4	9.520	8.102	2173468	1043109	40.802	38.536
45)	L9 AR-1268-5	9.945	8.391	16688924	8041001	38.007	40.630



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100923\
Data File : PO098612.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 05:28
Operator : YP/AJ
Sample : 04699-07
Misc : AR1268 LOD 40 PPB
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 06:00:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.480 min
Delta R.T.: -0.005 min
Response: 40932758
Conc: 21.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.635 min
Delta R.T.: -0.002 min
Response: 16145178
Conc: 20.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.294 min
Delta R.T.: 0.000 min
Response: 31732414
Conc: 26.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.642 min
Delta R.T.: -0.010 min
Response: 13667488
Conc: 25.51 ng/ml



#3 AR-1016-1


R.T.: 5.648 min
Delta R.T.: -0.010 min
Response: 166000
Conc: 3.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.735 min
Delta R.T.: 0.012 min
Response: 89592
Conc: 3.75 ng/ml




#4 AR-1016-2

R.T.: 5.702 min
Delta R.T.: 0.021 min
Response: 183753
Conc: 2.26 ng/ml



#4 AR-1016-2


R.T.: 4.735 min
Delta R.T.: -0.007 min
Response: 89592
Conc: 2.63 ng/ml



#5 AR-1016-3


R.T.: 5.732 min
Delta R.T.: -0.011 min
Response: 10528
Conc: 0.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.912 min
Delta R.T.: -0.006 min
Response: 35732
Conc: 1.99 ng/ml




#6 AR-1016-4

R.T.: 5.847 min
Delta R.T.: 0.005 min
Response: 12086
Conc: 0.31 ng/ml



#6 AR-1016-4


R.T.: 4.968 min
Delta R.T.: 0.008 min
Response: 47915
Conc: 3.20 ng/ml



#7 AR-1016-5


R.T.: 6.131 min
Delta R.T.: -0.007 min
Response: 17863
Conc: 0.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.195 min
Delta R.T.: 0.022 min
Response: 86847
Conc: 4.47 ng/ml




#8 AR-1221-1

R.T.: 4.701 min
Delta R.T.: 0.006 min
Response: 43463
Conc: 1.93 ng/ml



#8 AR-1221-1


R.T.: 3.861 min
Delta R.T.: 0.008 min
Response: 58260
Conc: 6.34 ng/ml



#9 AR-1221-2


R.T.: 4.789 min
Delta R.T.: 0.008 min
Response: 793031
Conc: 47.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.934 min
Delta R.T.: -0.005 min
Response: 258986
Conc: 38.95 ng/ml




#10 AR-1221-3

R.T.: 4.863 min
Delta R.T.: 0.005 min
Response: 45966
Conc: 0.94 ng/ml



#10 AR-1221-3


R.T.: 4.009 min
Delta R.T.: -0.005 min
Response: 45584
Conc: 2.26 ng/ml



#11 AR-1232-1


R.T.: 4.863 min
Delta R.T.: 0.005 min
Response: 45966
Conc: 1.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.009 min
Delta R.T.: -0.006 min
Response: 45584
Conc: 2.71 ng/ml




#12 AR-1232-2

R.T.: 5.386 min
Delta R.T.: -0.003 min
Response: 21040
Conc: 0.97 ng/ml



#12 AR-1232-2


R.T.: 4.735 min
Delta R.T.: -0.007 min
Response: 89592
Conc: 5.96 ng/ml



#13 AR-1232-3


R.T.: 5.702 min
Delta R.T.: 0.021 min
Response: 183753
Conc: 5.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.912 min
Delta R.T.: -0.006 min
Response: 35732
Conc: 4.57 ng/ml




#14 AR-1232-4

R.T.: 5.847 min
Delta R.T.: 0.004 min
Response: 12086
Conc: 0.71 ng/ml



#14 AR-1232-4


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 26865
Conc: 3.57 ng/ml



#15 AR-1232-5


R.T.: 5.930 min
Delta R.T.: -0.003 min
Response: 30338
Conc: 2.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.195 min
Delta R.T.: 0.021 min
Response: 86847
Conc: 10.74 ng/ml




#16 AR-1242-1

R.T.: 5.648 min
Delta R.T.: -0.012 min
Response: 166000
Conc: 3.65 ng/ml



#16 AR-1242-1


R.T.: 4.735 min
Delta R.T.: 0.012 min
Response: 89592
Conc: 4.28 ng/ml



#17 AR-1242-2


R.T.: 5.702 min
Delta R.T.: 0.021 min
Response: 183753
Conc: 2.68 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.735 min
Delta R.T.: -0.007 min
Response: 89592
Conc: 3.08 ng/ml




#18 AR-1242-3

R.T.: 5.732 min
Delta R.T.: -0.013 min
Response: 10528
Conc: 0.25 ng/ml



#18 AR-1242-3


R.T.: 4.912 min
Delta R.T.: -0.006 min
Response: 35732
Conc: 2.29 ng/ml



#19 AR-1242-4


R.T.: 5.847 min
Delta R.T.: 0.003 min
Response: 12086
Conc: 0.37 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 26865
Conc: 1.62 ng/ml




#20 AR-1242-5

R.T.: 6.567 min
Delta R.T.: -0.017 min
Response: 92920
Conc: 2.62 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T. : 5.526 min
Response: 0
Conc: N.D.



#21 AR-1248-1


R.T.: 5.648 min
Delta R.T.: -0.015 min
Response: 166000
Conc: 4.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.735 min
Delta R.T.: 0.013 min
Response: 89592
Conc: 5.72 ng/ml




#22 AR-1248-2

R.T.: 5.940 min
Delta R.T.: 0.003 min
Response: 25452
Conc: 0.49 ng/ml



#22 AR-1248-2


R.T.: 4.968 min
Delta R.T.: 0.009 min
Response: 47915
Conc: 2.18 ng/ml



#23 AR-1248-3


R.T.: 6.150 min
Delta R.T.: 0.008 min
Response: 39194
Conc: 0.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.002 min
Delta R.T.: 0.001 min
Response: 26865
Conc: 1.14 ng/ml




#24 AR-1248-4

R.T.: 6.551 min
Delta R.T.: 0.002 min
Response: 110621
Conc: 1.95 ng/ml



#24 AR-1248-4


R.T.: 5.195 min
Delta R.T.: 0.022 min
Response: 86847
Conc: 3.26 ng/ml



#25 AR-1248-5


R.T.: 6.606 min
Delta R.T.: 0.019 min
Response: 383006
Conc: 6.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.566 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.516 min
Delta R.T.: -0.005 min
Response: 61281
Conc: 0.97 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.525 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 6.740 min
Delta R.T.: -0.002 min
Response: 109533
Conc: 1.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.638 min
Delta R.T.: -0.036 min
Response: 508271
Conc: 14.76 ng/ml




#28 AR-1254-3

R.T.: 7.107 min
Delta R.T.: -0.002 min
Response: 160987
Conc: 1.71 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 6.078 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.404 min
Delta R.T.: 0.009 min
Response: 82621
Conc: 1.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.278 min
Delta R.T.: -0.028 min
Response: 171271
Conc: 6.24 ng/ml




#30 AR-1254-5

R.T.: 7.800 min
Delta R.T.: -0.015 min
Response: 263762
Conc: 3.78 ng/ml



#30 AR-1254-5


R.T.: 6.696 min
Delta R.T.: -0.029 min
Response: 268300
Conc: 6.19 ng/ml



#31 AR-1260-1


R.T.: 7.266 min
Delta R.T.: -0.002 min
Response: 14927
Conc: 0.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.218 min
Delta R.T.: 0.009 min
Response: 427717
Conc: 11.66 ng/ml




#32 AR-1260-2

R.T.: 7.539 min
Delta R.T.: 0.015 min
Response: 617731
Conc: 7.50 ng/ml



#32 AR-1260-2


R.T.: 6.379 min
Delta R.T.: -0.018 min
Response: 131841
Conc: 3.14 ng/ml



#33 AR-1260-3


R.T.: 7.885 min
Delta R.T.: 0.000 min
Response: 1421444
Conc: 22.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.561 min
Delta R.T.: 0.010 min
Response: 391877
Conc: 9.79 ng/ml




#34 AR-1260-4

R.T.: 8.103 min
Delta R.T.: -0.010 min
Response: 444182
Conc: 6.59 ng/ml



#34 AR-1260-4


R.T.: 7.014 min
Delta R.T.: -0.010 min
Response: 979276
Conc: 30.41 ng/ml



#35 AR-1260-5


R.T.: 8.437 min
Delta R.T.: -0.001 min
Response: 2306382
Conc: 17.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.259 min
Delta R.T.: -0.007 min
Response: 507931
Conc: 7.25 ng/ml




#36 AR-1262-1

R.T.: 7.885 min
Delta R.T.: -0.005 min
Response: 1421444
Conc: 16.34 ng/ml



#36 AR-1262-1


R.T.: 6.758 min
Delta R.T.: -0.005 min
Response: 796593
Conc: 17.60 ng/ml



#37 AR-1262-2


R.T.: 8.437 min
Delta R.T.: -0.006 min
Response: 2306382
Conc: 16.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.014 min
Delta R.T.: -0.008 min
Response: 979276
Conc: 24.29 ng/ml




#38 AR-1262-3

R.T.: 8.753 min
Delta R.T.: -0.009 min
Response: 7025661
Conc: 71.16 ng/ml



#38 AR-1262-3


R.T.: 7.542 min
Delta R.T.: -0.006 min
Response: 3804036
Conc: 122.54 ng/ml



#39 AR-1262-4


R.T.: 8.850 min
Delta R.T.: -0.001 min
Response: 6307862
Conc: 82.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.607 min
Delta R.T.: -0.005 min
Response: 3314045
Conc: 61.17 ng/ml




#40 AR-1262-5

R.T.: 9.520 min
Delta R.T.: -0.007 min
Response: 2173468
Conc: 44.81 ng/ml



#40 AR-1262-5


R.T.: 8.102 min
Delta R.T.: -0.005 min
Response: 1043109
Conc: 42.24 ng/ml



#41 AR-1268-1


R.T.: 8.753 min
Delta R.T.: -0.009 min
Response: 7025661
Conc: 40.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.542 min
Delta R.T.: -0.007 min
Response: 3804036
Conc: 41.63 ng/ml




#42 AR-1268-2

R.T.: 8.850 min
Delta R.T.: -0.007 min
Response: 6307862
Conc: 40.45 ng/ml



#42 AR-1268-2


R.T.: 7.607 min
Delta R.T.: -0.007 min
Response: 3314045
Conc: 42.74 ng/ml



#43 AR-1268-3


R.T.: 9.084 min
Delta R.T.: -0.008 min
Response: 5526444
Conc: 40.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.813 min
Delta R.T.: -0.007 min
Response: 2773414
Conc: 41.85 ng/ml




#44 AR-1268-4

R.T.: 9.520 min
Delta R.T.: -0.008 min
Response: 2173468
Conc: 40.80 ng/ml



#44 AR-1268-4


R.T.: 8.102 min
Delta R.T.: -0.007 min
Response: 1043109
Conc: 38.54 ng/ml



#45 AR-1268-5

R.T.: 9.945 min
Delta R.T.: -0.011 min
Response: 16688924
Conc: 38.01 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.391 min
Delta R.T.: -0.007 min
Response: 8041001
Conc: 40.63 ng/ml