

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100923\
 Data File : P0098612.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 05:28
 Operator : YP/AJ
 Sample : 04699-07
 Misc : AR1268 LOD 40 PPB
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2023

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/10/2023
 Supervised By :Ankita Jodhani 10/10/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 06:00:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.480	3.635	40932758	16145178	21.317	20.729
2) SA Decachlor...	10.294	8.642	31732414	14510978	26.142	27.085m
Target Compounds						
41) L9 AR-1268-1	8.753	7.542	7025661	3677468	40.460	40.244m
42) L9 AR-1268-2	8.850	7.607	6307862	3314045	40.454	42.741
43) L9 AR-1268-3	9.084	7.813	5526444	2773414	40.419	41.854
44) L9 AR-1268-4	9.520	8.102	2173468	1043109	40.802	38.536
45) L9 AR-1268-5	9.945	8.391	16688924	8041001	38.007	40.630

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100923\
Data File : PO098612.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 05:28
Operator : YP/AJ
Sample : 04699-07
Misc : AR1268 LOD 40 PPB
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT4-2023

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/10/2023
Supervised By :Ankita Jodhani 10/10/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 06:00:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

