

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100923\
 Data File : P0098570.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2023 10:04
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 03:17:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.482	3.634	101.4E6	38292037	52.786	49.164
2) SA Decachlor...	10.297	8.643	64506245	26385232	53.141	49.248m
Target Compounds						
21) L5 AR-1248-1	5.657	4.718	19051082	7840508	562.928	500.314
22) L5 AR-1248-2	5.932	4.954	27716750	11043733	535.188	501.835
23) L5 AR-1248-3	6.137	4.996	29821780	11641625	527.696	495.815
24) L5 AR-1248-4	6.543	5.168	31332850	13384530	552.720	501.777
25) L5 AR-1248-5	6.581	5.561	31417116	12292740	531.162	520.952



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100923\
Data File : P0098570.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2023 10:04
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 03:17:36 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.482 min
Delta R.T.: -0.003 min
Response: 101358794
Conc: 52.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.634 min
Delta R.T.: -0.004 min
Response: 38292037
Conc: 49.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.297 min
Delta R.T.: 0.003 min
Response: 64506245
Conc: 53.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.643 min
Delta R.T.: -0.009 min
Response: 26385232
Conc: 49.25 ng/ml m



#21 AR-1248-1


R.T.: 5.657 min
Delta R.T.: -0.005 min
Response: 19051082
Conc: 562.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.718 min
Delta R.T.: -0.004 min
Response: 7840508
Conc: 500.31 ng/ml




#22 AR-1248-2

R.T.: 5.932 min
Delta R.T.: -0.005 min
Response: 27716750
Conc: 535.19 ng/ml



#22 AR-1248-2


R.T.: 4.954 min
Delta R.T.: -0.004 min
Response: 11043733
Conc: 501.84 ng/ml



#23 AR-1248-3


R.T.: 6.137 min
Delta R.T.: -0.005 min
Response: 29821780
Conc: 527.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.996 min
Delta R.T.: -0.005 min
Response: 11641625
Conc: 495.82 ng/ml




#24 AR-1248-4

R.T.: 6.543 min
Delta R.T.: -0.006 min
Response: 31332850
Conc: 552.72 ng/ml



#24 AR-1248-4


R.T.: 5.168 min
Delta R.T.: -0.005 min
Response: 13384530
Conc: 501.78 ng/ml



#25 AR-1248-5

R.T.: 6.581 min
Delta R.T.: -0.006 min
Response: 31417116
Conc: 531.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.561 min
Delta R.T.: -0.005 min
Response: 12292740
Conc: 520.95 ng/ml