

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0100924\
 Data File : P0107015.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2024 09:30
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 02:34:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 04:48:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.375	3.646	616.7E6	211.2E6	51.411	52.952
2)	SA Decachlor...	10.056	8.645	127.7E6	137.9E6	51.028	49.131
Target Compounds							
16)	L4 AR-1242-1	5.523	4.729	135.8E6	47064409	534.146	527.939
17)	L4 AR-1242-2	5.546	4.748	193.2E6	65083797	529.614	531.924
18)	L4 AR-1242-3	5.608	4.924	122.2E6	35901760	535.285	534.357
19)	L4 AR-1242-4	5.704	5.007	97590369	36168050	533.873	531.521
20)	L4 AR-1242-5	6.437	5.530	79357345	42491937	520.048	528.235



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO100924\
Data File : PO107015.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2024 09:30
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 02:34:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 04:48:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.375 min
Delta R.T.: -0.001 min
Response: 616744865
Conc: 51.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: -0.001 min
Response: 211205503
Conc: 52.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.056 min
Delta R.T.: 0.000 min
Response: 127681885
Conc: 51.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.645 min
Delta R.T.: -0.002 min
Response: 137911768
Conc: 49.13 ng/ml



#16 AR-1242-1


R.T.: 5.523 min
Delta R.T.: -0.001 min
Response: 135787489
Conc: 534.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.729 min
Delta R.T.: -0.002 min
Response: 47064409
Conc: 527.94 ng/ml




#17 AR-1242-2

R.T.: 5.546 min
Delta R.T.: 0.000 min
Response: 193243416
Conc: 529.61 ng/ml



#17 AR-1242-2


R.T.: 4.748 min
Delta R.T.: -0.002 min
Response: 65083797
Conc: 531.92 ng/ml



#18 AR-1242-3


R.T.: 5.608 min
Delta R.T.: -0.001 min
Response: 122175689
Conc: 535.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.924 min
Delta R.T.: -0.002 min
Response: 35901760
Conc: 534.36 ng/ml




#19 AR-1242-4

R.T.: 5.704 min
Delta R.T.: -0.002 min
Response: 97590369
Conc: 533.87 ng/ml



#19 AR-1242-4


R.T.: 5.007 min
Delta R.T.: -0.001 min
Response: 36168050
Conc: 531.52 ng/ml



#20 AR-1242-5

R.T.: 6.437 min
Delta R.T.: 0.000 min
Response: 79357345
Conc: 520.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.530 min
Delta R.T.: 0.000 min
Response: 42491937
Conc: 528.24 ng/ml